



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Rappaport Vassiliadis Salmonella Enrichment Broth	
Product Code: BL017H	Expiry Date: 15/10/2019
Lot Number: 06361018	Release Date: 16/10/2018
Analytical Report Number: QAD/AR-18/09-1005	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Colour of Medium	Greenish blue colored Solution	Greenish blue colored Solution	Complies
pH Range	5.0 to 5.4	5.2	Complies
Quantity of Medium	10 ml	10 ml	Complies
Sterilization	Autoclaving at 10 lbs pressure (115 °C) for 15 minutes.	Autoclaving at 10 lbs pressure (115 °C) for 15 minutes.	Complies
Sterility Check	No growth should be observed after incubation at 30-35 °C for 24 - 48 hrs.	No growth observed after incubation at 30-35 °C for 24 - 48 hrs.	Complies

CULTURAL PARAMETERS:

Organisms	Growth Control Parameters			Recovery on XLD Agar (DM297H)		Remarks
	Incubation Temperature	Incubation time	Inoculum (CFU)	Growth	Color of colony	
<i>Salmonella typhimurium</i> ATCC 14028	30 – 35 °C	18 - 24 hrs.	84	Luxuriant	Red with black centres	Complies
<i>Salmonella abony</i> NCTC 6017	30 – 35 °C	18 - 24 hrs.	84	Luxuriant	Red with black centres	Complies
<i>Staphylococcus aureus</i> ATCC 6538	30 – 35 °C	18 - 24 hrs.	>10 ³	Inhibited	-	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Pooja Sutar) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

BL017HCOA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd
Manufacturer and Exporter
Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road,Thane (W)-400607,Maharashtra,India
[Tel:- +91-22-2589 5505](tel:+91-22-25895505)sales@micromasterlab.com