






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Sterile Disposable Plates-90mm	
Product Code: CRA019	Expiry Date: 18.04.2023
Lot Number: G0540418	Release Date: 19.04.2018
Analytical Report Number: QAD/AR-18/04-371	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Optically clear, Polystyrene, Petri plates, double packed in stacks of 10. Size: 90mm diameter x 15mm	Optically clear, Polystyrene, Petri plates, double packed in stacks of 10. Size: 90mm diameter x 15mm	Complies
Sterilization	The above batch of Petri plates was sterilized by Microtrol Sterilization Services Pvt. Ltd. by treatment with Ethylene oxide to the following parameters: Concentration of Ethylene oxide : NLT 750 gm/M3 Exposure time : 6 hours Humidity : Min. 50% RH Temperature : 50 Deg C	Concentration of Ethylene oxide: 750 gm/M3	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report