






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Sterile Disposable Plates-90mm</b>	
<b>Product Code: CRA019-100PC</b>	<b>Expiry Date: 18/04/2023</b>
<b>Lot Number: G0540418</b>	<b>Release Date: 19/04/2018</b>
<b>Analytical Report Number: QAD/AR-18/04-371</b>	<b>Page No: 1 of 1</b>

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Optically clear, Polystyrene, Petri plates, double packed in stacks of 10. Size: 90mm diameter x 15mm	Optically clear, Polystyrene, Petri plates, double packed in stacks of 10. Size: 90mm diameter x 15mm	Complies
Sterilization	The above batch of Petri plates was sterilized by Microtrol Sterilization Services Pvt. Ltd. by treatment with Ethylene oxide to the following parameters: Concentration of Ethylene oxide : NLT 750 gm/M3 Exposure time : 6 hours Humidity : Min. 50% RH Temperature : 50 Deg C	Concentration of Ethylene oxide: 750 gm/M3	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY (Madhura Kadam) MICROBIOLOGIST</b>	<b>CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY</b>	<b>APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE</b>

End Of The Report