CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Steri-Master Marker-Black	
Product Code: CRS002BKG	**Expiry Date: 05/12/2019**
Lot Number: 09761217	**Release Date: 15/01/2018**
Analytical Report Number: QAD/AR-18/01-04 A	**Page No: 1 of 1**

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Polypropylene, Individually packed, Black permanent, quick drying, alcohol resistant ink	Polypropylene, Individually packed, Black permanent, quick drying, alcohol resistant ink	Complies
Sterilization	The above batch of Clean Room Marker was sterilized by ISOMED (BRIT) by irradiation to the following parameters: Dose of irradiation 10-25 kgy.	Dose of irradiation 15Kgy	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY **(Madhura Kadam)** **MICROBIOLOGIST**	**CHECKED BY** **(Aditi Modak)** **HEAD MICROBIOLOGY**	**APPROVED BY** **(Rupali Sabale)** **HEAD QUALITY ASSURANCE**

End Of The Report

Micromaster Laboratories Pvt. Ltd

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall, Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com