

# PRODUCT SPECIFICATION SHEET



## Bile Esculin Agar (DM037)

### Intended Use

Bile Esculin Agar (DM037) is recommended for differential isolation and presumptive identification of group D *Streptococci* in food and pharmaceutical products.

### Product Summary and Explanation

Molecular taxonomic studies of the genus *Streptococcus* have placed enterococci, previously described as group D streptococci, in the genus *Enterococcus*.<sup>(1)</sup> Swan compared the use of an esculin medium containing 40% bile salts with the Lancefield serological method of grouping,<sup>(2)</sup> and reported that a positive reaction on the bile esculin medium correlated with a serological group D precipitin reaction. Bile Esculin Agar was originally formulated by Swan<sup>(2)</sup> for the isolation and identification of Group D *Streptococci* from food. Facklam and Moody<sup>(3, 4)</sup> further reported that using Bile Esculin Agar, Group D *Streptococci* could be differentiated from non Group D *Streptococci*. Group D *Streptococci* possess the group D lipoteichoic acid antigen in their cell walls. Former Group D species, which are predominant normal inhabitants of the human gastrointestinal tract, are termed as faecal *Streptococci* or *Enterococci*.<sup>(5)</sup> The unique ability of *Enterococci* to split esculin was reported by Meyer and Schonfeld.<sup>(6)</sup> *Enterococci* and Group D *Streptococci* hydrolyse esculin to esculetin and dextrose, which reacts with ferric citrate producing brownish black precipitate.<sup>(7)</sup> The use of esculin hydrolysis in identification of *Enterococci* was first cited by Rochaix.<sup>(8)</sup> Bile Esculin Agar was also shown to aid differentiation of *Enterobacteriaceae*, *Klebsiella*, *Enterobacter*, *Serratia* from other *Enterobacteriaceae* genera<sup>(9)</sup> on the basis of esculin hydrolysis. However, other tests such as salt tolerance should be performed for identifying *Enterococci*.<sup>(10)</sup>

### Principles of the Procedure

Bile Esculin Agar is highly nutritious. Peptic digest of animal tissue and beef extract serves as sources of carbon, nitrogen, amino acids, vitamins and essential growth nutrients. Oxgall is used to inhibit most of the other accompanying bacteria. Esculin in the medium is hydrolyzed to esculetin and dextrose. Esculetin reacts with ferric citrate to form a dark brown or black complex, visualized as a zone of black precipitate around the colonies. In slants, a positive reaction is indicated by blackening of more than half of the slant within 24-48 hours. If blackening is totally absent or if less than half of the slant is blackened within 24-48 hours, the test is negative. Viridans *Streptococci* sometimes exhibit a weak positive reaction. Also, *Leuconostoc*, *Pediococcus*, *Lactococcus* species causing human infections give a positive bile esculin test.

### Formula / Liter

Ingredients	Gms / Liter
Peptic digest of animal tissue	5.00
Beef extract	3.00
Oxgall	40.00
Esculin	1.00
Ferric citrate	0.50
Agar	15.00
Final pH: 6.6 ± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	

### Precautions

1. For Laboratory Use only.
2. IRRITANT. Irritating to eyes, respiratory system, and skin.

### Directions

1. Suspend 64.5 grams of the medium in one liter of distilled water/purified water.
2. Heat to boiling to dissolve the medium completely.
3. Mix well and dispense into tubes or flasks as desired.
4. Autoclave at 121°C, 15 psi pressure, for 15 minutes / validated cycle.
5. Allow the tubed medium to solidify in slanted position.



# PRODUCT SPECIFICATION SHEET



## Quality Control Specifications

Dehydrated Appearance	Light yellow to brownish yellow homogeneous free flowing powder
Prepared Medium	Amber coloured, clear to slightly opalescent gel with a bluish tinge forms in Petri plates or in tubes as slants
Reaction of 6.5% Solution	pH : $6.6 \pm 0.2$ at $25^{\circ}\text{C}$
Gel Strength	Firm, comparable with 1.5% Agar gel

**Expected Cultural Response:** Cultural characteristics observed in an increased atmosphere of Carbon dioxide after an incubation at  $35-37^{\circ}\text{C}$  for 18-24 hours.

Sr. No.	Organisms	Results to be achieved			
		Inoculum (CFU)	Growth	Recovery	Esculin Hydrolysis
1.	<i>Enterococcus faecalis</i> ATCC 29212	50 -100	good-luxuriant	$\geq 50\%$	positive reaction, blackening of medium around the colony
2.	<i>Proteus mirabilis</i> ATCC 25933	50 -100	good-luxuriant	$\geq 50\%$	negative reaction
3.	<i>Streptococcus pyogenes</i> ATCC 19615	50-100	none-poor	$\leq 10\%$	negative reaction

The organisms listed are the minimum that should be used for quality control testing.



Bile Esculin Agar (DM037)  
1. *Enterococcus faecalis*  
2. Control



*Enterococcus faecalis* ATCC 29212

## Test Procedure

Refer to appropriate references for instructions on specific material being tested for group D streptococci.

## Results

- Any blackening around the colony of the plated medium indicates a positive result; if no blackening occurs, the test is negative.
- For slants, if more than half of the slant is blackened within 24-48 hours, the test is positive; if less than half is blackened or no blackening occurs within 24-48 hours, the test is negative.

## Storage

Store the sealed bottle containing the dehydrated medium at  $10 - 30^{\circ}\text{C}$ . Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.



# PRODUCT SPECIFICATION SHEET



## Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

## Limitations of the Procedure

1. Consult appropriate texts for detailed information and recommended procedures.
2. Strains of *Lactococcus*, *Leuconostoc* and *Pediococcus* that give a positive bile-esculin reaction have been isolated from human infections.
3. Occasional strains of viridans streptococci blacken the medium or display weakly positive reactions.

## Packaging

**Product Name :** Bile Esculin Agar.

**Product Code :** DM037

**Available Pack sizes :** 100gm / 500gm

## References

1. Schleifer, K. H., and R. Kilpper-Balz. 1987. Molecular and chemotaxonomic approaches to the classification of streptococci, enterococci and lactococci: a review. Syst. Appl. Microbiol. 10:1-19.
2. Swan, A. 1954. The use of bile-esculin medium and of Maxted's technique of Lancefield grouping in the identification of enterococci (group D streptococci). J. Clin. Pathol. 7:160.
3. Facklam R., 1972, Appl. Microbiol., 23:1131.
4. Facklam R. R and Moody M. D., 1970, Appl. Microbiol., 20(2):245.
5. Koneman E. W., Allen S. D., Janda W. M., Schreckenberger P. C., Winn W. C. Jr., 1992, Colour Atlas and Textbook of Diagnostic Microbiology, 4th Ed., J. B. Lippincott Company
6. Meyer and Schonfeld, 1926, Zentralbl. Bakteriologie, Parasitenkunde, Infektionskrankheiten, Hygiene, Abt. Origin. 99:402.
7. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. I, Williams and Wilkins, Baltimore.
8. Rochaix, 1924, Compt. Rend. Soc. Biol., 90:771.
9. Facklam R., 1973, Appl. Microbiol., 26:138. Swan, 1954, J. Clin. Pathol., 7:160.
10. Murray P. R., Baron E. J., Tenover J. C., Tenover F. C., 2003, Manual of Clinical Microbiology, ASM, Washington, D.C.

## Further Information

For further information please contact your local MICROMASTER Representative.



**MICROMASTER LABORATORIES PRIVATE LIMITED**

DM037PSS, QAD/FR/024, Rev.00

Unit 38/39, Kalpataru Industrial Estate,

Off G.B. Road, Near 'R-Mall', Thane (W) - 400607. M.S. INDIA.

Ph: +91-9320126789/9833630009/9819991103

Email: [sales@micromasterlab.com](mailto:sales@micromasterlab.com)

## Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

