

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Brilliant Green Bile Agar	
Product Code: DM045	Expiry Date : 11/05/2022
Lot Number: 02900417	Release Date: 11/05/2017
Analytical Report Number: QAD/AR-17/05-555	Page No: 1 of 2

DEHYDRATED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color	Pinkish purple to light purple	Pinkish purple to light purple	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies
PREPARED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color of medium	Bluish purple coloured	Bluish purple coloured	Complies
Clarity of medium (After sterilization)	Slightly opalescent gel forms in petri plates	Slightly opalescent gel formed in petri plates	Complies
pH (After sterilization)	6.70-7.10	7.10	Complies
Gel strength	Firm as compared to 1.0% Agar gel.	Firm as compared to 1.0% Agar gel.	Complies

CULTURAL PARAMETERS:

Organisms	Incubation time/ temp	No. of organisms inoculated	No. of organisms recovered	Percentage recovery	Growth	Colour of colony	Remarks
<i>Escherichia coli</i> ATCC 25922	35-37°C / 18-24 hrs.	76	58	76%	Good-luxuriant	Deep red (may Have bile Precipitate)	Complies
<i>Enterobacter aerogenes</i> ATCC 13048	35-37°C / 18-24 hrs.	77	57	78%	Good-luxuriant	Pink	Complies
<i>Salmonella Enteritidis</i> ATCC 13076	35-37°C / 18-24 hrs.	80	60	77%	Good-luxuriant	Colourless to Light pink	Complies
<i>Staphylococcus aureus</i> ATCC 25923	35-37°C / 18-24 hrs.	>=10 ⁸	0	0%	Inhibited	NA	Complies

Rev. 00


10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Brilliant Green Bile Agar	
Product Code: DM045	Expiry Date : 11/05/2022
Lot Number: 02900417	Release Date: 11/05/2017
Analytical Report Number: QAD/AR-17/05-555	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

