



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Czapek Dox Broth</b>	
<b>Product Code: DM070</b>	<b>Expiry Date : 16/02/2023</b>
<b>Lot Number: 00790218</b>	<b>Release Date: 17/02/2018</b>
<b>Analytical Report Number: QAD/AR-18/02-136</b>	<b>Page No: 1 of 2</b>

<b>DEHYDRATED MEDIUM</b>			
<b>TEST CONDUCTED</b>	<b>SPECIFICATION</b>	<b>OBSERVATIONS</b>	<b>REMARKS</b>
Color	Cream to yellow	Cream to yellow	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies
<b>PREPARED MEDIUM</b>			
<b>TEST CONDUCTED</b>	<b>SPECIFICATION</b>	<b>OBSERVATIONS</b>	<b>REMARKS</b>
Color of medium	Light yellow coloured	Light yellow coloured	Complies
Clarity of medium (After sterilization)	Clear to slightly opalescent solution in tubes	Clear to slightly opalescent solution in tubes	Complies
pH (After sterilization)	7.10-7.50 at 25°C	7.12	Complies
Gel strength	Not Applicable	Not Applicable	Complies

### **CULTURAL PARAMETERS:**

<b>Organisms</b>	<b>Incubation time/ temp</b>	<b>No. of organisms inoculated</b>	<b>Growth</b>	<b>Remarks</b>
<i>Aspergillus brasiliensis</i> ATCC 16404	25-30°C / 48-72 hrs.	82	Good-Luxuriant	Complies
<i>Candida albicans</i> ATCC 10231	25-30°C / 48-72 hrs.	75	Good-Luxuriant	Complies
<i>Saccharomyces cerevisiae</i> ATCC 9763	25-30°C / 48-72 hrs.	81	Good-Luxuriant	Complies






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Czapek Dox Broth	
Product Code: DM070	Expiry Date : 16/02/2023
Lot Number: 00790218	Release Date: 17/02/2018
Analytical Report Number: QAD/AR-18/02-136	Page No: 2 of 2

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY (Deepti Nair) MICROBIOLOGIST</b>	<b>CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY</b>	<b>APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE</b>

End Of The Report