



## PRODUCT SPECIFICATION SHEET

### Endo Agar (DM099)

#### Intended Use

Endo Agar (DM099) is a selective medium recommended for confirmation of the presumptive test for members of the coliform group.

#### Product Summary and Explanation

The majority of media for enterobacteriaceae, developed in the early years, utilized either mixtures of bile salts or individual bile salts as selective agents to achieve inhibition of gram-positive species. In 1904, Endo reported the development of a culture medium for the differentiation of lactose fermenters from the non-fermenters in which no bile salts were used<sup>(1)</sup>. Inhibition of gram-positive microorganisms was achieved by the sodium sulfite and basic fuchsin contained in the formulation. Endo's Fuchsin Sulphite Infusion Agar was the original name for this medium<sup>(2)</sup>, which is known today as Endo Agar. It was developed initially in order to facilitate the isolation and identification of the typhoid bacillus. The original formula has been modified extensively since its introduction. Many modifications of this media have been done over the years. Endo Agar is recommended by APHA as an important medium in the microbiological examination of water and wastewater, dairy products and foods<sup>(3-5)</sup>. Endo Agar is used to confirm the detection and enumeration of coliform bacteria following presumptive test of drinking water. It is also used for the detection and isolation of coliforms and fecal coliforms from milk, dairy products and food.

#### Principles of the Procedure

The medium contains peptic digest of animal tissue which provides nitrogen, carbon, vitamins and minerals required for bacterial growth. The selectivity of Endo Agar is due to the sodium sulfite/basic fuchsin combination, which results in the suppression of gram-positive microorganisms. It is classified as only slightly selective since other media contain more potent inhibitors of the gram-positive microorganisms. Coliforms ferment the lactose, produce pink to rose-red colonies and similar coloration of the medium. The colonies of organisms that do not ferment lactose are colorless to faint colonies against the pink background of the medium. With *Escherichia coli*, this reaction is very pronounced as the fuchsin crystallizes, exhibiting a permanent greenish metallic luster to the colonies. Medium should be stored away from light to avoid photo-oxidation

#### Formula / Liter

Ingredients	Gms / Litre
Peptic digest of animal tissue	10.00
Lactose	10.00
Dipotassium phosphate	3.50
Sodium sulphite	2.50
Basic fuchsin	0.50
Agar	15.00
Final pH: 7.5 ± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	





## PRODUCT SPECIFICATION SHEET

### Precautions

1. For Laboratory Use only.
2. TOXIC. Harmful if swallowed, inhaled, or absorbed through skin. May cause allergic reaction and breathing difficulties to sensitive individuals. May cause irritation to skin, eyes, and respiratory tract. Possible carcinogen.

### Directions

1. Suspend 41.5 g of the medium in one liter of deionised water.
2. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes.
3. Mix well before pouring into sterile Petri plates.
4. If the solidified culture medium is somewhat too red, then to remove the colour add a few drops (max. 1 ml/litre) of a freshly prepared 10% Sodium sulphite solution and boil.

### Quality Control Specifications

<b>Dehydrated Appearance</b>	Light pink to purple, homogeneous free flowing powder
<b>Prepared Medium</b>	Orangish pink coloured, clear to slightly opalescent gel with fine precipitate forms in petri-plates
<b>Reaction of 4.15% Solution</b>	pH 7.5 ± 0.2 at 25°C
<b>Gel Strength</b>	Firm, compared to 1.5% Agar Gel.

**Expected Cultural Response:** Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Sr. No.	Organisms	Results to be achieved			
		Inoculum	Growth	Recovery	Colour of Colony
1.	<i>Bacillus subtilis</i> ATCC 6633	>=10 <sup>3</sup>	Inhibited	0%	--
2.	<i>Enterobacter aerogenes</i> ATCC 13048	50-100	Good-luxuriant	>=50%	Pink
3.	<i>Enterococcus faecalis</i> ATCC 29212	50-100	None-poor	<=10%	Pink, small
4.	<i>Escherichia coli</i> ATCC 25922	50-100	Good-luxuriant	>=50%	Pink to rose red with metallic sheen
5.	<i>Klebsiella pneumoniae</i> ATCC 13883	50-100	Good-luxuriant	>=50%	Pink, mucoid
6.	<i>Proteus vulgaris</i> ATCC 13315	50-100	Good-luxuriant	>=50%	Colourless to pale pink
7.	<i>Pseudomonas aeruginosa</i> ATCC 27853	50-100	Good-luxuriant	>=50%	Colourless, irregular
8.	<i>Salmonella typhi</i> ATCC 6539	50-100	Good-luxuriant	>=50%	Colourless to pale pink
9.	<i>Shigella sonnei</i> ATCC 25931	50-100	Good-luxuriant	>=50%	Colourless to pale pink
10.	<i>Staphylococcus aureus</i> ATCC 25923	>=10 <sup>3</sup>	Inhibited	0%	--





## PRODUCT SPECIFICATION SHEET

11.	<i>Enterobacter cloacae</i> ATCC 13047	50-100	Good	40-50%	Pink
12.	<i>Salmonella typhimurium</i> ATCC 14028	50-100	Good-luxuriant	>=50%	Colourless
13.	<i>Salmonella enteritidis</i> ATCC 13076	50-100	Good-luxuriant	>=50%	Colourless
14.	<i>Shigella flexneri</i> ATCC 12022	50-100	Good-luxuriant	>=50%	Colourless

The organisms listed are the minimum that should be used for quality control testing.



Characteristic metallic colonies of *E. coli*.



Lactose positive colonies of *Salmonella* spp &  
Lactose negative colonies of *Enterobacter* spp

### Test Procedure

1. Use standard procedures to obtain isolated colonies from specimens.
2. A nonselective medium should also be streaked to increase the chance of recovery when the population of gram negative organisms is low and to provide an indication of other organisms present in the specimen.
3. Incubate plates, protected from light, at  $35 \pm 2^\circ\text{C}$  for 18-24 hours.
4. If negative after 24 hours, reincubate an additional 24 hours.

### Results

Following incubation, examine the plates for presence of colored colonies. All colonies that are pink and have the characteristic metallic sheen are considered coliforms. The sheen may cover the entire colony, may only be in the center or may appear only around the edges.





## PRODUCT SPECIFICATION SHEET

---

### Storage

Store the sealed bottle containing the dehydrated medium at 10 - 30°C. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.

### Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

### Limitations of the Procedure

1. If the inoculum is too heavy, the sheen may be suppressed.
2. Occasionally, non-coliform organisms may produce typical sheen colonies. Coliform organisms may also occasionally produce atypical colonies, including dark red or nucleated colonies without sheen.

### Packaging

**Product Name: Endo Agar (DM099)**

**Product Code : DM099**

**Available Pack sizes : 100gm / 500gm**

### References

1. Endo. 1904. Zentralbl. Bakteriol., Abt. 1, Orig. 35:109.
2. Levin and Schoenlein. 1930. A compilation of culture media for the cultivation of microorganisms. Williams & Wilkins, Baltimore, Md.
3. Eaton, Rice and Baird (ed). 2005. Standard methods for the examination of water and wastewater, 21st ed., online. American Public Health Association, Washington, D.C.
4. Wehr and Frank (ed.). 2004. Standard methods for the examination of dairy products. 17th ed. American Public Health Association, Washington, D.C.
5. Downes and Ito (ed.). 2001. Compendium of methods for the microbiological examination of foods, 4th ed. American Public Health Association, Washington, D.C.

### Further Information

For further information please contact your local MICROMASTER Representative.



**MICROMASTER LABORATORIES PRIVATE LIMITED**

Unit 38/39, Kalpataru Industrial Estate,  
Off G.B. Road, Near 'R-Mall' , Thane (W) - 400607. M.S. INDIA.  
Ph: +91-22-25895505, 4760, 4681. Cell: 9320126789.

Email: [micromaster@micromasterlab.com](mailto:micromaster@micromasterlab.com)  
[sales@micromasterlab.com](mailto:sales@micromasterlab.com)

DM099PSS,QAD/FR/024,Rev.00/01.01.2018





## PRODUCT SPECIFICATION SHEET

---

	Checked By	Approved By
<i>Adelam</i> 01.01.2018	<i>Rusdak</i> 01.01.2018	<i>Lobole</i> 01.01.2018





## PRODUCT SPECIFICATION SHEET

---

### Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

