# PRODUCT SPECIFICATION SHEET



# Mannitol Salt Broth (DM161)

# Intended Use

Mannitol Salt Broth (DM161) is recommended for selective isolation of presumptive pathogenic Staphylococci.

# Product Summary and Explanation

In 1942, Koch,<sup>(1)</sup> reported that only *Staphylococci* grow on agar media containing 7.5% sodium chloride. Chapman<sup>(2)</sup> further studied this phenomenon in greater detail and concluded that the addition of 7.5% sodium chloride to phenol red mannitol agar results in an improved medium for the isolation of plasma-coagulating *Staphylococci*. Mannitol Salt Broth is prepared as suggested by Chapman<sup>(2)</sup> and is used for the selective isolation of pathogenic *Staphylococci*. This medium is recommended for the detection and enumeration of coagulase-positive *Staphylococci* in milk<sup>(3)</sup> food<sup>(4)</sup> and other specimens. Mannitol Salt Broth is used for the isolation of presumptive pathogenic *Staphylococci*.

## Principles of the Procedure

Mannitol Salt Broth contains beef extract and proteose peptone which makes it very nutritious as they provide essential growth factors and trace nutrients. High sodium chloride (7.5%) content inhibits many other bacteria except *Staphylococci*. Mannitol is the fermentable carbohydrate source. The differential action of the medium is attributed to D-Mannitol. Phenol red is a pH indicator. The change in the colour of the medium is due to the reactivity of phenol red to the pH of the medium; phenol red at pH 8.4 and yellow at 6.8.

## Formula / Liter

Ingredients	Gms / Liter		
Proteose peptone	10.00		
Beef extract	1.00		
Sodiumchloride	75.00		
D-Manni tol	10.00		
Phenol red	0.025		
Final pH: 7.4 ± 0.2 at 25°C			
Formula may be adjusted and/or supplemented as required to meet performance			

Precautions

specifications

- 1. For Laboratory Use only.
- 2. IRRITANT. Irritating to eyes, respiratory system, and skin.
- 3. This product contains 7.5% sodium chloride as one of its ingredients. On repeated exposure to air and absorption moisture sodium chloride has tendency to form lumps, therefore we strongly recommend storage in tightly closed containers in dry place away from bright light.

## Directions

- 1. Suspend 96.02 grams of the medium in one liter of distilled water.
- 2. Heat if necessary to dissolve the medium completely.
- 3. Dispense as desired.
- 4. Autoclave at 121°C, 15 psi pressure, for 15 minutes / validated cycle.

## Quality Control Specifications

Dehydrated Appearance	Light yellow to pink homogeneous free flowing powder	
Prepared Medium	epared Medium Red coloured clear solution in tubes	
Reaction of 9.6% Solution	pH:7.4±0.2 at 25°C	
Gel Strength	Not Applicable	

## Expected Cultural Response: Cultural characteristics observed after an incubation at 35-37°C for 18-48 hours.

Sr. Organisms Results to be achieved
--------------------------------------





# PRODUCT SPECIFICATION SHEET

No.		Inoculum (CFU)	Growth	Colour of Medium
1.	Escherichia coli ATCC 25922	>=10 <sup>3</sup>	inhibited	
2.	Staphylococcus aureus ATCC 25923	50 - 100	good-luxuriant	yellow
3.	Staphylococcus epidermidisATCC 12228	50 - 100	fair-good	red

The organisms listed are the minimum that should be used for quality control testing.

# Test Procedure

Refer to appropriate references for standard test procedures.

# Results

- 1. Staphylococcus aureus ferments mannitol to produce yellow coloured medium.
- 2. Most coagulase-negative species of Staphylococci and Micrococci do not ferment mannitol and therefore the medium remains red in colour.
- 3. Presumptive Staphylococcus showing yellow coloured medium should be further tested for production of coagulase.

## Storage

Store the sealed bottle containing the dehydrated medium at 10 - 30°C. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.

## Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

## Limitations of the Procedure

1. A possible S. aureus must be confirmed by the coagulase test.

- 2. Also the organism should be subcultured to a less inhibitory medium not containing excess salt to avoid the possible interference of salt with coagulase testing or other diagnostic tests (e.g. Nutrient Broth) (MOO2).
- 3. Few strains of *S. aureus* may exhibit delayed mannitol fermentation.
- 4. Negative results should therefore be re-incubated for an additional 24 hours before being discarded.
- 5. Consult appropriate texts for detailed information and recommended procedures.

# Packaging

Product Name : Mannitol Salt Broth Product Code : DM161 Available Pack sizes : 500gm

## References

- 1. Koch. 1942. Zentralbl. Bakteriol. Parasitenkd. Abt. I Orig. 149:122.
- 2. Chapman G.H., 1945, J. Bact., 50:201.
- 3. Marshall R. (Ed.), 1992, Standard Methods for the Examination of Dairy Products, 16th ed., APHA, Washington, D.C.
- 4. Bacteriological Analytical Manual, 1995, Food and Drug Administration, 8th ed., AOAC, International, U.S.A.

# Further Information

For further information please contact your local MICROMASTER Representative.



#### MICROMASTER LABORATORIES PRIVATE LIMITED

DM161PS5,QAD/FR/024,Rev.00

Unit 38/39, Kalpataru Industrial Estate,

Off G.B. Road, Near 'R-Mall', Thane (W) - 400607. M.S. INDIA. Ph: +91-9320126789/9833630009/9819991103 Email: sales@micromasterlab.com





# PRODUCT SPECIFICATION SHEET

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

