# PRODUCT SPECIFICATION SHEET



## Motility Test Medium (DM171)

#### Intended Use

Motility Test Medium (DM171) is recommended for detection of bacterial motility.

#### Product Summary and Explanation

Bacterial motility can be observed directly on microscopic slide or it can be visualized on motility media having agar concentration of 0.4% or less.<sup>(1)</sup> In 1936, Tittsler and Sandholzer reported on the use of semisolid agar for the detection of bacterial motility.<sup>(2)</sup> Motility Test Medium is a modification of their formulation. Certain species of motile bacteria will show diffuse growth throughout the entire medium, while others may show diffusion from one or two points appearing as nodular outgrowths along the stab. Tittsler and Sandholzer reported tubes incubated for one day gave identical results with the hanging drop method, and incubation for two days permitted demonstration of motility in an additional 4% of cultures tested. Motility can be visualized as a diffused zone of growth flaring outfrom the line of inoculation.<sup>(1)</sup> Hanging-drop technique in motility tests has practical difficulties, which is efficiently eliminated by use of culture-based methods using semi-solid media, as in semisolid media; the results obtained are macroscopic and cumulative.

#### Principles of the Procedure

Motility Test Medium contains tryptose which acts as a source of essential growth nutrients required for bacterial metabolism. Sodium chloride maintains the osmotic equilibrium of the medium. Small amount of agar helps to create a semisolid medium.

#### Formula / Liter

Ingredients	Gms / Litre			
Tryptose	10.00			
Sodiumchloride	5.00			
Agar	5.00			
Final pH: 7.2 ± 0.2 at 25°C				
Formula may be adjusted and/or supplemented as required to meet performance				
specifications				

#### Precautions

- 1. For Laboratory Use only.
- 2. IRRITANT, mainly irritating to eyes, respiratory system, and skin.

#### Directions

- 1. Suspend 20 grams of the medium in one liter of distilled water/purified water.
- 2. Heat to boiling to dissolve the medium completely.
- 3. Dispense into tubes and sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes.
- 4. Allow tubed medium to cool in an upright position.

#### Quality Control Specifications

Dehydrated Appearance	Cream to yellow homogeneous free flowing powder	
Prepared Medium	Light yellow coloured clear to slightly opalescent gel forms in tubes as butts	
Reaction of 2.0% Solution	pH 7.2 <u>+</u> 0.2 at 25°C	
Gel Strength	Semisolid, comparable with 0.5% Agar gel	

#### Expected Cultural Response: Cultural Cultural characteristics observed after an incubation at 35 - 37°C for 18 - 48 hours.

Sr.	Organisms	Results to be achieved		
No.		Inoculum (CFU)	Growth	Colour of colony
1.	Escherichia coli ATCC 25922	50-100	good-luxuriant	positive, growth away from stabline causing turbidity
2.	Enterobacter aerogenes ATCC 13048	50-100	good- luxuriant	positive, growth away from stabline





# PRODUCT SPECIFICATION SHEET

				causing turbidity
3.	Klebsiella pneumoniae ATCC 13883	50-100	good- luxuriant	negative, growth along the stabline, surrounding medium remains clear
4.	Salmonella Enteritidis ATCC 13076	50-100	good- luxuriant	positive, growth away from stabline causing turbidity
5.	Staphylococcus aureus ATCC 25923	50-100	good-luxuriant	negative, growth along the stabline, surrounding medium remains clear

The organisms listed are the minimum that should be used for quality control testing.

### Test Procedure

- 1. Inoculate tubes with a pure culture by stabbing the centre of the column of medium to greater than half the depth.
- 2. Incubate tubes for 24-48 hours at  $35 \pm 2^{\circ}C$  in an aerobic atmosphere.

#### Results

- 1. After incubation, observe the tubes for growth in relation to the stab line. Non-motile organisms grow only along the line of inoculation, while motile organisms spread out from the line of inoculation and may even grow throughout the medium.
- 2. Negative tubes can be re-incubated at 25 ± 2°C for an additional 5 days, if desired.
- 3. All weak or equivocal motility results should be confirmed by flagellum stain or by direct wet microscopy (hanging drop).<sup>(3,4)</sup>
- 4. Refer to appropriate references for interpretation of results.

### Storage

Store the sealed bottle containing the dehydrated medium at 10 - 30°C. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.

#### Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

### Limitations of the Procedure

- 1. Many organisms fail to grow deep in semisolid media; inoculating pour plates may be advantageous.
- 2. Consult appropriate texts for detailed information and recommended procedures.

#### Packaging

#### Product Name : Motility Test Medium Product Code : DM171 Available Pack sizes : 100gm / 500gm References

- 1. Koneman E. W., Allen S. D., Janda W. M., Schreckenberger P. C., Winn W. C. Jr., (Eds.), 1992, Colour Atlas and Textbook of Diagnostic Microbiology, 4th Ed., J. B. Lippinccott Company.
- 2. Tittsler R.P. and Sandholzer L.A., 1936, J. Bacteriol., 31:575.
- 3. DAmato R. F., and Tomfohrede K. M., 1981, J. Clin. Microbiol., 14 (3), 347-348.
- 4. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore.

### Further Information

For further information please contactyour local MICROMASTER Representative.



#### MICROMASTER LABORATORIES PRIVATE LIMITED

DM171PSS,QAD/FR/024,Rev.00

Unit 38/39, Kalpataru Industrial Estate, Off G.B. Road, Near'R-Mall', Thane (W) - 400607. M.S. INDIA. Ph: +91-9320126789/9833630009/9819991103 Email: <u>sales@micromasterlab.com</u>





# PRODUCT SPECIFICATION SHEET

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

