



PRODUCT SPECIFICATION SHEET

Peptone Water (DM192)

Intended Use

Peptone Water (DM192) is an all purpose growth medium and may be used as a base for carbohydrate fermentation media.

Product Summary and Explanation

Peptone Water may be used as a growth medium or as a basis of carbohydrate fermentation media. A pure culture in Peptone water is a convenient inoculum for a series of fermentation tubes or other diagnostic media. The formulation of Peptone water makes it useful for cultivating non-fastidious organisms.⁽¹⁾ Peptone Water, adjusted to pH 8.4, is suitable for the cultivation and enrichment of *Vibrio cholerae* from infected material.⁽²⁾ The medium was formerly used for the performance of the indole test, but now better results can be obtained by use of Tryptone water. Peptone Water may be modified to make it suitable for carbohydrate fermentation tests by the addition of Andrade indicator and the required carbohydrate.

Principles of the Procedure

Peptic digest of animal tissue provides essential nutrients. Sodium Chloride maintains the osmotic balance of the medium. If desired for carbohydrate fermentation studies Durham's tube may be used to detect the gas production if produced.

Formula / Liter

Ingredients	Gms / Litre
Peptic digest of animal tissue	10.00
Sodium chloride	5.00
Final pH: 7.2± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	

Precautions

1. For Laboratory Use only.
2. IRRITANT. Irritating to eyes, respiratory system, and skin.

Directions

1. Suspend 15 grams of the medium in one liter of distilled water.
2. Add the test carbohydrate in desired quantity and dissolve completely.
3. Dispense in tubes with or without inverted Durham's tubes.
4. Autoclave at 121°C, 15 psi pressure, for 15 minutes / validated cycle.

Quality Control Specifications

Dehydrated Appearance	Cream to yellow colored, homogeneous, free flowing powder
Prepared Medium	Light amber colored clear solution without any precipitate
Reaction of 1.5% Solution	pH 7.2± 0.2 at 25°C
Gel Strength	Not Applicable

Expected Cultural Response: Cultural response on Peptone Water observed after incubation at 35-37°C for 18-24 hours

Sr. No.	Organisms	Results to be achieved		
		Inoculum (CFU)	Growth	Indole test
1.	<i>Escherichia coli</i> ATCC 25922	50 -100	Luxuriant	Positive reaction, red ring at the interface of the medium on addition of Kovac's reagent (IR002)





PRODUCT SPECIFICATION SHEET

2.	<i>Salmonella typhimurium</i> ATCC 14028	50 -100	Luxuriant	Negative reaction, no red ring at the interface of the medium on addition of Kovac's reagent (IR002)
3.	<i>Staphylococcus aureus</i> ATCC 25923	50-100	Luxuriant	Negative reaction, no red ring at the interface of the medium on addition of Kovac's reagent (IR002)

The organisms listed are the minimum that should be used for quality control testing.

Test Procedure

For Performing Carbohydrate Fermentation

1. Inoculate tubes with test organism.
2. Incubate tubes at $35 \pm 2^{\circ}\text{C}$ for 18-48 hours.
3. Observe for color change.

For Performing the Indole Test

1. Using aseptic technique, suspend an Indole Test Strip 10mm above the surface of a 24 or 48 hour culture.
2. Incubate at 37°C for 5-30 minutes.
3. Alternatively inoculate peptone water with the test organism. Incubate at 37°C for 18-24 hours.
4. Add few drops of indole reagent to the culture broth.
5. Observe for the formation red ring.


Results

For determining Carbohydrate Fermentation Patterns

Acid is produced when carbohydrates are fermented. This is indicated by a yellow color in the medium. Gas production is indicated by the presence of gas bubbles in the fermentation tube.

For performing Indole Test

Observe for the formation of violet color on the strip which indicates a positive test for indole production or observe for the red ring at the interface of the medium.



Peptone Water (DM192)

1. *Escherichia coli* ATCC 25922-Indole positive
2. *Staphylococcus aureus* ATCC 25923-Indole negative

Storage

Store the sealed bottle containing the dehydrated medium at $10 - 30^{\circ}\text{C}$. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.





PRODUCT SPECIFICATION SHEET

Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

Limitations of the Procedure

1. Medium is pink in color when hot but becomes colorless upon cooling.
2. *Vibrio spp.* should not be incubated longer than 18-20 hours. Longer incubation may cause the development of suppressed forms.⁽³⁾

Packaging

Product Name : Peptone Water

Product Code : DM192

Available Pack sizes : 100gm / 500gm

References

1. MacFaddin, J.F. 1985. Media for isolation-cultivation-identification-maintenance of medical bacteria, vol,1, p. 610-612. Williams & Wilkins, Baltimore, MD.
2. Cruickshank R. (1968) 'Medical Microbiology' 11th ed., Livingstone Ltd., London, p. 268.
3. Finegold, S.M., and W. Martin. 1982. Bailey and Scott's diagnostic microbiology, 6th ed. St. Louis.

Further Information

For further information please contact your local MICROMASTER Representative.



MICROMASTER LABORATORIES PRIVATE LIMITED

DM192PSS,QAD/FR/024,Rev.00/01.01.2018




Unit 38/39, Kalpataru Industrial Estate,

Off G.B. Road, Near 'R-Mall', Thane (W) - 400607. M.S. INDIA.

Ph: +91-22-25895505, 4760, 4681. Cell: 9320126789.

Email: micromaster@micromasterlab.com

sales@micromasterlab.com

Prepared By	Checked By	Approved By
 01.01.2018	 01.01.2018	 01.01.2018
Microbiologist	Head Quality Control	Head Quality Assurance





PRODUCT SPECIFICATION SHEET

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

