



PRODUCT SPECIFICATION SHEET

Casein Soyabean Digest Agar (Soyabean Casein Digest Agar) (Tryptone Soya Agar) (DM247)

Intended Use

Casein Soyabean Digest Agar (Soyabean Casein Digest Agar) (Tryptone Soya Agar) (DM247) is recommended for cultivation of a wide variety of microorganisms.

Product Summary and Explanation

In 1955, Leavitt et al.⁽¹⁾ discovered that Soyabean Casein Digest Agar facilitated vigorous growth of aerobic and anaerobic microorganisms. Soyabean Casein Digest Agar is a widely used medium, which supports the growth of wide variety of organisms even that of fastidious ones such as *Neisseria*, *Listeria*, and *Brucella* etc. The medium is used in *USP* Growth Promotion testing and when testing the suitability of counting methods in the presence of product.⁽²⁾ TSA has a multitude of uses in the clinical laboratory including maintenance of stock cultures, plate counting, isolation of microorganisms from a variety of specimen types and as a base for media containing blood.⁽³⁻⁶⁾ It is also recommended for use in industrial applications when testing water and wastewater,⁽⁷⁾ food,⁽⁸⁻¹³⁾ dairy products,⁽¹⁴⁾ and cosmetics.^(9,15) The medium with addition of blood provides perfectly defined haemolysis zones, while preventing the lysis of erythrocytes due to its sodium chloride content. It has been frequently used in the health industry to produce antigens, toxins etc. It's simple and inhibitor-free composition makes it suitable for the detection of antimicrobial agents in the food and other products. Tryptone Soya Agar is recommended by various pharmacopoeias as sterility testing medium.^(2,16) Tryptone Soya Agar conforms as per USP⁽²⁾ and is used in microbial limit test and antimicrobial preservative - effective test. Gunn et al⁽¹⁷⁾ used this medium for the growth of fastidious organisms and study of haemolytic reaction after addition of 5% v/v blood. Soyabean Casein Digest Agar does not contains X and V growth factors. It can be conveniently used in determining the requirements of these growth factors by isolates of *Haemophilus* by the addition of X-factor (ID007), V-factor (ID008), and X+V factor discs (ID009) factor to inoculated TSA plates.⁽⁴⁾

Principles of the Procedure

Soyabean Casein Digest Agar contains pancreatic digest of casein and papaic digest of soyabean which provides amino acids, long chain peptides and essential nutrients required for the growth of microorganisms. Sodium chloride maintains the osmotic balance. *Haemophilus* species may be differentiated by their requirements for X and V factors. Paper strips impregnated with these factors are placed on the surface of the medium after inoculation with the test organism. Following incubation, a zone of growth around the strip indicates a requirement for the factor(s).

Formula / Liter

Ingredients	Gms / Liter
Pancreatic digest of casein	15.00
Papaic digest of soyabean (soyabean)	5.00
Sodium chloride	5.00
Agar	15.00
Final pH: 7.3 ± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	

Precautions

1. For Laboratory Use only.
2. IRRITANT. Irritating to eyes, respiratory system, and skin.





PRODUCT SPECIFICATION SHEET

Directions

1. Suspend 40 grams of the medium in one liter of distilled water.
2. Heat to boiling, to dissolve the medium completely.
3. Autoclave at 121°C, 15 psi pressure, for 15 minutes / validated cycle.
4. Mix well and pour into sterile petri plates.

Quality Control Specifications

Dehydrated Appearance	Cream to yellow homogeneous free flowing powder
Prepared Medium	Basal Medium : Light yellow coloured clear to slightly opalescent gel. After addition of 5-7%w/v sterile defibrinated blood : Cherry red coloured opaque gel forms in Petri plates
Reaction of 4.0% Solution	pH : 7.3 ± 0.2 at 25°C
Gel Strength	Firm, comparable with 1.5% Agar gel

Expected Cultural Response: Cultural characteristics was observed after an incubation for Bacterial at 30-35°C 18-24 hours and for Fungal at 20-25°C <=5days

Sr. No.	Organisms	Results to be achieved (CFU)					
		Inoculum (CFU)	Observed Lot value (CFU)	Recovery	Observed Lot value (CFU) w/blood	Recovery w/blood	Heamolysis
1.	<i>Bacillus subtilis</i> ATCC 6633	50 - 100	35 -100	≥70 %	35 -100	≥70 %	none
2.	<i>Staphylococcus aureus</i> ATCC 25923	50 - 100	35 -100	≥70 %	35 -100	≥70 %	beta
3.	<i>Staphylococcus aureus</i> ATCC 6538	50 - 100	35 -100	≥70 %	35 -100	≥70 %	beta
4.	<i>Escherichia coli</i> ATCC 25922	50 - 100	35 -100	≥70 %	35 -100	≥70 %	none
5.	<i>Escherichia coli</i> ATCC 8739	50 - 100	35 -100	≥70 %	35 -100	≥70 %	none
6.	<i>Escherichia coli</i> NCTC 9002	50 - 100	35 -100	≥70 %	35 -100	≥70 %	none
7.	<i>Pseudomonas aeruginosa</i> ATCC 27853	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
8.	<i>Pseudomonas aeruginosa</i> ATCC 9027	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
9.	<i>Salmonella</i> Abony NCTC 6017	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
10.	<i>Micrococcus luteus</i> ATCC 9341	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
11.	<i>Streptococcus pneumonia</i> ATCC 6305	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
12.	<i>Salmonella</i> Typhimurium ATCC 14028	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
13.	<i>Candida albicans</i> ATCC 10231	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
14.	<i>Candida albicans</i> ATCC 2091	50 - 100	35 -100	≥70 %	35 -100	≥70 %	-
15.	<i>Aspergillus brasiliensis</i> ATCC 16404	50 - 100	25 -70	50-70%			

The organisms listed are the minimum that should be used for quality control testing.





PRODUCT SPECIFICATION SHEET

Test Procedure

1. For clinical specimens, refer to appropriate standard references for details on testing protocol to obtain isolated colonies from specimens.
2. For water, food, dairy or cosmetic samples, refer to appropriate standard references for details on test methods.
3. For pharmaceutical samples, refer to *USP General Chapter <61>* for details on the examination of nonsterile products and performing microbial enumeration tests.

Results

After incubation, it is desirable to have isolated colonies of organisms from the original sample. Subculture colonies of interest, so that positive identification can be made by means of biochemical and/or serological testing.

Storage

Store the sealed bottle containing the dehydrated medium at 10 - 30°C. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.

Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

Limitations of the Procedure

1. For identification, organisms must be in pure culture. Morphological, biochemical and/or serological tests should be performed for final identification.
2. Consult appropriate texts for detailed information and recommended procedures.

Packaging

Product Name : Casein Soyabean Digest Agar (Soyabean Casein Digest Agar) (Tryptone Soya Agar)

Product Code : DM247

Available Pack sizes : 100gm/ 500gm

References

1. Leavitt, J. M., I. J. Naidorf and P. Shugaevsky. 1955. The undetected anaerobe in endodontics: a sensitive medium for detection of both aerobes and anaerobes. *The NY J. Dentist*. 25:377-382.
2. United States Pharmacopeial Convention, Inc. 2008. *The United States pharmacopeia 31/The national formulary 26*, Supp. 1, 8-1-08, online. United States Pharmacopeial Convention, Inc., Rockville, Md.
3. MacFaddin. 1985. *Media for isolation-cultivation-identification-maintenance of medical bacteria*, vol. 1. Williams & Wilkins, Baltimore, Md.
4. Forbes, Sahn and Weissfeld. 2007. *Bailey & Scott's diagnostic microbiology*, 12th ed. Mosby Inc., St. Louis, Mo.
5. Murray, Baron, Jorgensen, Landry and Pfaller (ed.). 2007. *Manual of clinical microbiology*, 9th ed. American Society for Microbiology, Washington, D.C.
6. Isenberg and Garcia (ed.). 2004 (update, 2007). *Clinical microbiology procedures handbook*, 2nd ed. American Society for Microbiology, Washington, D.C.
7. Eaton, Rice and Baird (ed.). 2005. *Standard methods for the examination of water and wastewater*, 21st ed., online. American Public Health Association, Washington, D.C.
8. Downes and Ito (ed.). 2001. *Compendium of methods for the microbiological examination of foods*, 4th ed. American Public Health Association, Washington, D.C.
9. U.S. Food and Drug Administration. 2001. *Bacteriological analytical manual*, online. AOAC International, Gaithersburg, Md.
10. U.S. Department of Agriculture. *Microbiology laboratory guidebook*, online. Food Safety and Inspection Service, USDA, Washington, D.C.
11. Horwitz (ed.). 2007. *Official methods of analysis of AOAC International*, 18th ed., online. AOAC International, Gaithersburg, Md.





PRODUCT SPECIFICATION SHEET

12. Health Canada. The compendium of analytical methods, online. Food Directorate, Health Products and food Branch, Health Canada, Ottawa, Ontario Canada.
13. International Organization for Standardization. 1994. Microbiology - General guidance for the detection of presumptive pathogenic *Yersinia enterocolitica*. ISO 10273, 1st ed., 1994-12-15. International Organization for Standardization, Geneva, Switzerland.
14. Wehr and Frank (ed.). 2004. Standard methods for the examination of dairy products, 17th ed. American Public Health Association, Washington, D.C.
15. Curry, Joyce and McEwen. 1993. CTFA microbiology guidelines. The Cosmetic, Toiletry and Fragrance Association, Inc., Washington, D.C.
16. Indian Pharmacopoeia, 2007, Govt. of India, Ministry of Health and Family Welfare, New Delhi, India.
17. Gunn B. A., Ohashi D K., Gaydos C. A., Holt E. S., 1977, J. Clin. Microbiol., 5(6) : 650.

Further Information

For further information please contact your local MICROMASTER Representative.



MICROMASTER LABORATORIES PRIVATE LIMITED

DM247PSS, QAD/FR/024, Rev.00/01.01.2018

Unit 38/39, Kalpataru Industrial Estate,
 Off G.B. Road, Near 'R-Mall' , Thane (W) - 400607. M.S. INDIA.
 Ph: +91-22-25895505, 4760, 4681. Cell: 9320126789.
 Email: micromaster@micromasterlab.com
sales@micromasterlab.com

Prepared By	Checked By	Approved By
Microbiologist	Head Quality Control	Head Quality Assurance

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise





PRODUCT SPECIFICATION SHEET

specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

