

PRODUCT SPECIFICATION SHEET

TCBS Agar (Thiosulphate Citrate Bile Sucrose Agar) (DM253)

Intended Use

TCBS Agar (Thiosulfate-Citrate-Bile-Sucrose Agar) is a selective differential medium for isolating and cultivating *Vibrio cholerae* and other enteropathogenic *Vibrio* species from clinical specimens and other materials.

Product Summary and Explanation

Vibrio spp. are most widely recognized for their role in causing Cholera and diarrhea diseases and also causing food poisoning. The isolation and cultivation of *Vibrio* species has been enhanced by the development of media which are highly selective for vibrios. It was developed by Kobayashi et al.⁽¹⁾, who modified the formula of Nakanishi.⁽²⁾, which resulted in TCBS Agar, a selective and differential media for the isolation and cultivation of vibrios. Although this medium was originally designed for the isolation of *V. cholerae* and *V. parahaemolyticus*, most *Vibrios* grow to healthy large colonies with many different colonial morphologies. Enrichment in Alkaline Peptone Water, followed by isolation on TCBS Agar is routinely used for isolation of *V. cholera* and other *Vibrio* spp.⁽¹⁻⁴⁾ TCBS Agar complies with the recommendations of the World Health Organization WHO and APHA for the selective isolation of *V. cholerae* and *V. parahaemolyticus*.^(6,7,8) This medium is recommended for isolating *Vibrio* spp. from stool specimens^(1,2,6) and is mentioned in Standard Methods for food testing.^(7,8) It is highly selective, meets the nutritional requirements of *Vibrio* spp., and allows vibrios to compete with intestinal flora. All members of the genus are able to grow in media containing increased salt concentrations and some species are halophilic.⁽⁶⁾

Principles of the Procedure

Enzymatic digest of casein and Peptic digest of animal tissue supplies the nitrogenous source and amino acids. Yeast extract provides the vitamin-B complex. TCBS has a very high pH (8.5-9.5) which suppresses growth of intestinal flora other than *Vibrio* spp.⁽⁹⁾. Ox bile, a derivative of bile salts and sodium citrate inhibit gram-positive bacteria and coliforms⁽¹⁰⁾. One percent sodium chloride is incorporated into the medium to provide optimum growth and metabolic activity of halophilic *Vibrio* spp. Sodium thiosulphate serves as a good source of sulphur, which in combination with ferric citrate detects the production of hydrogen sulphide. Sucrose is fermentable sugar, with the help of bromothymol blue and thymol blue indicators, allows for the differentiation of those *Vibrio* spp. which utilize sucrose. The alkaline pH of the medium improves the recovery of *V. cholerae*. Strains of *V. cholerae* produce yellow colonies on TCBS Agar because of fermentation of sucrose. *V. alginolyticus* also produce yellow colonies. *V. parahaemolyticus* is a sucrose non-fermenting organism and therefore produces blue-green colonies, as does *V. vulnificus*. species that are sucrose-fermenters may form yellow colonies⁽⁹⁾. TCBS Agar is not a suitable medium for oxidase testing of *Vibrio* species⁽¹⁰⁾. A few strains of *V. cholerae* may appear green or colourless on TCBS Agar due to delayed sucrose fermentation⁽⁹⁾.

Formula / Liter

Ingredients	Gms / Litre
Proteose peptone	10.00
Yeast Extract	5.00
Sodium Citrate	10.00
Sodium Thiosulfate	10.00
Ox bile	8.00
Sucrose	20.00
Sodium Chloride	10.00
Ferric Citrate	1.00
Bromothymol Blue	0.04
Thymol Blue	0.04
Agar	15.00
Final pH: 8.6 ± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	

PRODUCT SPECIFICATION SHEET

Precautions

1. For Laboratory Use only.
2. IRRITANT, mainly irritating to eyes, respiratory system, and skin. Handle in accordance with good laboratory hygiene and safety practice. Wash hands before breaks and at the end of workday. To protect, use safety glasses and gloves during handling.
3. Use personal protective equipment. Avoid dust formation. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Avoid breathing dust.
4. Do not let product enter drains.
5. After use, prepared plates, specimen containers and other contaminated materials must be sterilized by autoclaving before discarding.

Directions

1. Suspend 89.08 g of the medium in one liter of distilled water.
2. Heat with frequent agitation and boil completely to dissolve the medium.
3. DO NOT OVERHEAT. DO NOT AUTOCLAVE.
4. Cool to 50°C and pour into sterile Petri plates.

Quality Control Specifications

Dehydrated Appearance	Light yellow to light tan colored, homogeneous, free flowing powder
Solution	8.9% Solution in Distilled or deionized water is soluble on boiling, bluish green / forest green colored, and very slightly to slightly opalescent.
Prepared Medium	Bluish / Forest green, clear to slightly opalescent gel forms in petri plates.
Reaction of 8.9% Solution	pH 8.6 ± 0.2 at 25°C
Gel Strength	Firm, compared to 1.5% Agar Gel.

Expected Cultural Response: Cultural response on TCBS Agar observed after an incubation at 35-37°C for 18-48 hours.

Sr. No.	Organisms	Results to be achieved			
		Inoculum (CFU)	Growth	Recovery %	Colour of colony
1.	<i>Enterococcus faecalis</i> ATCC 29212	$\geq 10^3$	Inhibited	0%	-
2.	<i>Escherichia coli</i> ATCC 25922	$\geq 10^3$	Inhibited	0%	-
3.	<i>Shigella flexneri</i> ATCC 12022	$\geq 10^3$	Inhibited	0%	-
4.	<i>Vibrio cholerae</i> ATCC 15748	50-100	Good luxuriant	≥ 50 %	Yellow
5.	<i>Vibrio parahaemolyticus</i> ATCC 17802	50-100	Good luxuriant	≥ 50 %	Bluish Green
6.	<i>Vibrio fluvialis</i> ATCC 33809	50-100	Good luxuriant	≥ 50 %	Yellow
7.	<i>Vibrio vulnificus</i> ATCC 29306	50-100	Fair - good	≥ 30 %	Greenish Yellow

The organisms listed are the minimum that should be used for quality control testing.

Test Procedure

1. Observe aseptic techniques.
2. The agar surface should be smooth and moist, but without excessive moisture.
3. Streak the specimen as soon as possible after it is received in the laboratory. Specimens such as rectal swabs, feces, vomitus, fish or food samples may be swabbed directly onto the plated medium.
4. Heavy inoculation is recommended, especially if specimens are not fresh, as the medium is highly selective and vibrios tend to die rather easily.
5. The medium should be inoculated heavily with faecal specimens because growth of few species may be inhibited on the medium due to fermentation of sucrose and accumulation of acids.
6. Swabs containing specimen material should be transported to the laboratory in Cary and Blair Transport Medium^(3,5,17) if a delay in reaching the laboratory is anticipated.
7. Specimens for cultivation of vibrios should not be frozen.
8. Incubate plates, protected from light, at 35 ± 2°C in an aerobic atmosphere for 18-24 hours.


PRODUCT SPECIFICATION SHEET

Results


1. After 18 - 24 hours of incubation at 32-35°C, sucrose-fermentating vibrios (*V. cholerae*, *V. alginolyticus*, *V. hareyi*, *V. cincinnatiensis*, *V. fluvialis*, *V. furnissii*, *V. metschnikovii*, and some *V. vulnificus*) appear smooth, opaque, thin-edged yellow colonies on TCBS Agar.⁽³⁾
2. Most other clinically important vibrios, including *V. parahaemolyticus*, do not ferment sucrose and appear as bluish green colonies.⁽²⁾
3. Additional biochemical and/or biochemical tests are necessary for a final identification and for a differentiation of sucrose-fermenting and sucrose-nonfermenting species.^(1,2,7,8)
4. TCBS Agar is highly selective for *Vibrio* species. However, occasional isolates of *Pseudomonas* and *Aeromonas* may also form blue green colonies on TCBS Agar.⁽⁹⁾
5. Any H₂S negative colony of TCBS Agar can be considered presumptive positive for *Vibrio*.

Typical colonial morphology on TCBS Agar is as follows:

<i>V. cholera</i>	Yellow
<i>V. mimicus</i> , <i>V. parahaemolyticus</i> , <i>V. damsela</i>	Green
<i>V. vulnificus</i>	Green (85%) or yellow (15%)
<i>V. hollisae</i>	Green (very poor growth)
<i>V. fluvialis</i> , <i>V. furnissii</i> , <i>V. alginolyticus</i>	Yellow
<i>V. metschnikovii</i>	Yellow (reduced growth)
<i>Pseudomonas aeruginosa</i>	Inhibition, partial to complete; blue
<i>Proteus species</i>	Inhibition, partial to complete; yellow to translucent
<i>Escherichia coli</i>	Inhibition, partial to complete; translucent



V. cholera ferment sucrose, this results in a pH shift production of yellow colonies.



V. parahaemolyticus do not ferment sucrose and produce bluish green colonies.

Storage

Store the sealed bottle containing the dehydrated medium at 2 - 30°C. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.

Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

PRODUCT SPECIFICATION SHEET

Limitations of the Procedure

1. For identification, organisms must be in pure culture. Morphological, biochemical, and/or serological tests should be performed for final identification.
2. Colonies that appear yellow on TCBS Agar will produce unsatisfactory oxidase reactions.
3. Sucrose-fermenting *Proteus* spp. produce yellow colonies which may resemble those of *Vibrio*.⁽⁸⁾
4. Consult appropriate texts for detailed information and recommended procedures.^(10,11,13-16)
5. A single medium is rarely adequate for detecting all organisms of potential significance in a specimen. Cultures of specimens grown on selective media should, therefore, be compared with specimens cultured on nonselective media to obtain additional information and help ensure recovery of potential pathogens.

Packaging

Product Name : TCBS Agar / Thiosulfate Citrate Bile Salts Sucrose Agar

Product Code : DM253

Available Pack sizes : 100gm / 500gm

References

1. Kobayashi T., Enomoto S., Sakazaki R., and Kuwahara S., 1963, Jap. J. Bacteriol., 18: 387.
2. Nakanishi Y., 1963, Modern Media 9: 246.
3. Downes F. P. and Ito K., (Eds.), 2001, Compendium of Methods for the Microbiological Examination of Foods, 4th Ed., American Public Health Association, Washington, D.C.
4. Clesceri L. S., Greenberg A. E. and Eaton A. D., (Eds.), 1998, Standard Methods for the Examination of Water and Wastewater, 20th Ed., American Public Health Association, Washington, D.C.
5. Murray P. R., Baron J. H., Pfaller M. A., Jorgensen J. H. and Tenover F. C., (Eds.), 2003, Manual of Clinical Microbiology, 8th Ed., American Society for Microbiology, Washington, D.C.
6. Furniss A. L., Lee J. V. and Donovan T. J., 1978, The Vibrios, Public Health Laboratory Service Monograph Series No. 11, Maidstone Public Health Laboratory, H.M.S.O., London, England.
7. Forbes B. A., Sahm A. S. and Weissfeld D. F., 1998, Bailey & Scotts Diagnostic Microbiology, 10th Ed., Mosby, Inc. St. Louis, Mo.
8. Howard B., 1994, Clinical and Pathogenic Microbiology, 2nd Ed., The C.V. Mosby.
9. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. Williams & Wilkins, Baltimore, Md.
10. Morris G. K., Merson M. H., Huq A. K., Kibria A. K. and Black R., 1979, J. Clin. Microbiol., 9:79
11. Farmer, J.J., J.M. Janda, and K. Birkhead. 2003. *Vibrio*, p. 706-718. In P.R. Murray, E.J. Baron, J.H. Jorgensen, M.A. Pfaller, and R.H. Tenover (ed.), Manual of clinical microbiology, 8th ed. American Society for Microbiology, Washington, D.C.
12. Wacksmuth, I.K. 1984. Laboratory detection of enterotoxin, p. 93-115. In P. Ellner (ed.), Infectious diarrheal diseases: current concepts and laboratory procedures. Marcel Dekker, Inc., New York.
13. National Committee for Clinical Laboratory Standards. 2001. Approved Guideline M29-A2. Protection of laboratory workers from occupationally acquired infections, 2nd ed. NCCLS, Wayne, PA.
14. Garner, J.S. 1996. Hospital Infection Control Practices Advisory Committee, U.S. Department of Health and Human Services, Centers for Disease Control and Prevention. Guideline for isolation precautions in hospitals. Infect. Control Hospital Epidemiol. 17:53-80.
15. U.S. Department of Health and Human Services. 1999. Biosafety in microbiological and biomedical laboratories, HHS Publication (CDC), 4th ed. U.S. Government Printing Office, Washington, D.C.
16. Directive 2000/54/EC of the European Parliament and of the Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work (seventh individual directive within the meaning of Article 16(1) of Directive 89/391/EEC). Official Journal L262, 17/10/2000, p. 0021-0045.
17. Dewitt, W.E., E.J. Gangarosa, I. Huq, and A. Zarifi. 1971. Holding media for the transport of *Vibrio cholerae* from field to laboratory. Am. J. Trop. Med. Hyg. 20:685-688.
18. Holt, J.G., N.R. Krieg, P.H.A. Sneath, J.T. Staley, and S.T. Williams (ed.). 1994. Bergey's Manual. of determinative bacteriology, 9th ed. Williams & Wilkins, Baltimore.
19. MacFaddin, J.F. 2000. Biochemical tests for identification of medical bacteria, 3rd ed. Lippincott Williams & Wilkins, Baltimore.
20. Koneman, E.W., S.D. Allen, W.M. Janda, P.C. Schreckenberger, and W.C. Winn, Jr. 1997. Color atlas and textbook of diagnostic microbiology, 5th ed. Lippincott-Raven, Philadelphia.
21. Isenberg, H.D. (ed.). 2004. Clinical microbiology procedures handbook, vol. 1, 2 and 3, 2nd ed. American Society for Microbiology, Washington, D.C.

PRODUCT SPECIFICATION SHEET

Further Information

For further information please contact your local MICROMASTER Representative.



MICROMASTER LABORATORIES PRIVATE LIMITED

DM253PSS, QAD/FR/024, Rev.00

Unit 38/39, Kalpataru Industrial Estate,
Off G.B. Road, Near 'R-Mall', Thane (W) - 400607. M.S. INDIA.
Ph: +91-9320126789/9833630009/9819991103
Email: sales@micromasterlab.com

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.