



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Letheen Broth	
Product Code: DM315U	Expiry Date: 19/04/2022
Lot Number: 02210417	Release Date: 19/04/2017
Analytical Report Number: QAD/AR-17/04-431	Page No: 1 of 2

### DEHYDRATED MEDIUM

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color	Part A: Off-white to yellow colored powder Part B: Yellow colored solution	Part A: Off-white to yellow colored powder Part B: Yellow colored solution	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies

### PREPARED MEDIUM

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color of medium	Light yellow colored solution	Light yellow colored solution	Complies
Clarity of medium (after sterilization)	Clear to slightly opalescent solution (Mixture of Part A and Part B)	Clear to slightly opalescent solution (Mixture of Part A and Part B)	Complies
pH (after sterilization)	Auto pH		
Gel strength	Not Applicable	Not Applicable	Not Applicable

**CULTURAL PARAMETERS:** Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours.

Organisms	Incubation Time/ temp	No. of organisms inoculated	Growth	Remarks
<i>Staphylococcus aureus</i> ATCC 6538	35-37°C / 24-48 hrs	85	Good-luxuriant	Complies
<i>Escherichia coli</i> ATCC 8739	35-37°C / 24-48 hrs	88	Good-luxuriant	Complies

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com






# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Lethen Broth	
Product Code: DM315U	Expiry Date: 19/04/2022
Lot Number:02210417	Release Date: 19/04/2017
Analytical Report Number: QAD/AR-17/04-431	Page No: 2 of 2

## Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Daniya Kazi) MICROBIOLOGIST	<b>CHECKED BY</b> (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

**Micromaster Laboratories Pvt. Ltd.**  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

