



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Asparagine Proline Broth	
Product Code: DM318	Expiry Date: 25/05/2023
Lot Number: 03030518	Release Date: 26/05/2018
Analytical Report Number: QAD/AR-18/05-486	Page No: 1 of 2

DEHYDRATED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color	White to cream colored powder	White to cream colored powder	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies
PREPARED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color of medium	Colourless clear Solution	Colourless clear Solution	Complies
Clarity of medium (After sterilization)	Clear solution without any precipitate	Clear solution without any precipitate	Complies
pH (After sterilization)	7.10 to 7.50 at 25°C	7.20	Complies
Gel strength	Not Applicable	Not Applicable	Not Applicable

CULTURAL PARAMETERS:

Organisms	Incubation time/ temp	No. of organisms inoculated	Growth	Remarks
<i>Escherichia coli</i> ATCC25922	35-37°C / 24-48 hrs.	58	None to poor	Complies
<i>Pseudomonas aeruginosa</i> ATCC 27853	35-37°C / 24-48 hrs.	58	Luxuriant with greenish yellow pigment	Complies






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Asparagine Proline Broth	
Product Code: DM318	Expiry Date: 25/05/2023
Lot Number: 03030518	Release Date: 26/05/2018
Analytical Report Number: QAD/AR-18/05-486	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report