



PRODUCT SPECIFICATION SHEET

Tryptose Broth w/ Thiamine HCl (DM438)

Intended Use

Tryptose Broth w/ Thiamine HCl (DM438) is an infusion free medium for isolation, differentiation and cultivation of fastidious microorganisms.

Product Summary and Explanation

Tryptose media, prepared without extract or infusion of meat, are recommended for the cultivation and isolation of pathogenic and saprophytic bacteria.⁽¹⁾ Huddleson found tryptose media to be equal or superior to meat infusion media, providing uniformity for the cultivation and differentiation of fastidious organisms.^(2,3) McCullough et al. reported that addition of thiamine, dextrose and iron salts increased growth of *Brucella suis*.^(4,5) This medium can be used as general purpose medium for cultivation of wide variety of organisms. It can also be supplemented with defibrinated blood (sheep, horse) to prepare blood containing medium for the isolation of fastidious organisms like *Brucella*. Tryptose Broth with thiamine HCl is recommended by APHA⁽⁶⁾ and Diagnostic Procedures and Reagents⁽⁷⁾ for the isolation and cultivation of *Brucella* species and also *Streptococci*, *Meningococci*, *Pneumococci* and other pathogenic bacteria.⁽⁸⁾

Principles of the Procedure

Tryptose Broth w/ Thiamine HCl contains tryptose serves as nitrogen source. Dextrose is the energy source. Sodium chloride provides osmotic balance.

Formula / Liter

Ingredients	Gms / Liter
Tryptose	20.00
Dextrose	1.00
Sodium chloride	5.00
Thiamine hydrochloride	0.005
Final pH: 7.2 ± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	

Precautions

1. For Laboratory Use only.
2. IRRITANT. Irritating to eyes, respiratory system, and skin.
3. Biosafety Level 2 practices, containment equipment and facilities are recommended for activities with clinical specimens of human or animal origin containing or potentially containing pathogenic *Brucella* spp.
4. Biosafety Level 3 practices, containment equipment and facilities are recommended for all manipulations of cultures of the pathogenic *Brucella* spp. and for experimental animal studies.

Directions

1. Suspend 26 grams of the medium in one liter of distilled water.
2. If desired, add 0.5-1% agar to the medium.
3. Heat to boiling to dissolve the media completely.
4. Autoclave at 121°C, 15 psi pressure, for 15 minutes / validated cycle.
5. For blood media, aseptically add 5% v/v sterile defibrinated blood.
6. Mix well and dispense as desired.





PRODUCT SPECIFICATION SHEET

Quality Control Specifications

Dehydrated Appearance	Cream to yellow homogeneous free flowing powder
Prepared Medium	Basal Medium : Yellow coloured clear to slightly opalescent solution After addition of 5% v/v sterile defibrinated blood: cherry red coloured opaque solution forms in tubes
Reaction of 2.6% solution	pH 7.2 ± 0.2 at 25°C
Gel Strength	Not Applicable

Expected Cultural Response: Cultural characteristics observed after an incubation at 35-37°C for 48-72 hours with added 5% v/v sterile defibrinated blood in presence of 10% Carbon dioxide (CO₂).

Sr. No.	Organisms	Results to be achieved
		Growth
1.	<i>Brucella melitensis</i> ATCC 4309	good-luxuriant
2.	<i>Brucella suis</i> ATCC 4314	good-luxuriant
3.	<i>Streptococcus pneumoniae</i> ATCC 6303	good-luxuriant
4.	<i>Streptococcus pyogenes</i> ATCC 19615	good-luxuriant

The organisms listed are the minimum that should be used for quality control testing.

Test Procedure

Refer appropriate references for standard test procedures.

Results

Refer appropriate references and procedures for interpretation of results.

Storage

Store the sealed bottle containing the dehydrated medium at 10 - 30°C. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.

Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

Limitations of the Procedure

1. For identification, organisms must be in pure culture. Morphological, biochemical and/or serological tests should be performed for final identification.
2. Consult appropriate texts for detailed information and recommended procedures.

Packaging

Product Name : Tryptose Broth w/ Thiamine HCl

Product Code : DM438

Available Pack sizes : 500gm





PRODUCT SPECIFICATION SHEET

References

1. Huddleson I. F., 1943, Brucellosis in man and animals, rev., Ed., The Commonwealth Fund, New York, N.Y.
2. Ruiz Castañeda M., 1947, Proc. Soc. Exp. Biol. Med., 64:114.
3. Huddleson I. F., 1939, Brucellosis in Man and Animals, Oxford University Press, Oxford, England.
4. McCullough W. G., Mills R. L., Herbst E. J., Roessler W. J. and Brewer C. R., 1947, J. Bacteriol., 53:5.
5. Atlas R. M., 2004, Handbook of Microbiological Media, Lawrence C. Parks, (Ed.), 3rd Edition, CRC Press.
6. Standard Methods for the Microbiological Examination of Dairy Products, 9th Ed., 1948, APHA Inc., New York.
7. Diagnostic Procedures and Reagents, 1950, 3rd Edition, APHA, New York.
8. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore

Further Information

For further information please contact your local MICROMASTER Representative.



MICROMASTER LABORATORIES PRIVATE LIMITED

DM438PSS,QAD/FR/024,Rev.00/01.01.2018

Unit 38/39, Kalpataru Industrial Estate,
Off G.B. Road, Near 'R-Mall', Thane (W) - 400607. M.S. INDIA.
Ph: +91-22-25895505, 4760, 4681. Cell: 9320126789.

Email: micromaster@micromasterlab.com
sales@micromasterlab.com

Prepared By	Checked By	Approved By
Microbiologist	Head Quality Control	Head Quality Assurance

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

