



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Violet Red Bile Broth	
Product Code: DM440	Expiry Date : 16/06/2023
Lot Number: 03620618	Release Date: 16/06/2018
Analytical Report Number: QAD/AR-18/06-561	Page No: 1 of 2

DEHYDRATED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color	Light yellow to pinkish beige	Light yellow to pinkish beige	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies
PREPARED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color of medium	Reddish purple coloured	Reddish purple coloured	Complies
Clarity of medium (After sterilization)	Clear solution in tubes	Clear solution in tubes	Complies
pH (After sterilization)	7.20-7.60 at 25°C	7.54	Complies
Gel strength	NA	NA	NA

***DO NOT AUTOCLAVE.**



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY




Product Name: Violet Red Bile Broth	
Product Code: DM440	Expiry Date : 16/06/2023
Lot Number: 03620618	Release Date: 16/06/2018
Analytical Report Number: QAD/AR-18/06-561	Page No: 2 of 2

CULTURAL PARAMETERS:

Organisms	Incubation time/ temp	No. of organisms inoculated	Growth	Colour of medium	Remarks
<i>Escherichia coli</i> ATCC 25922	35-37°C / 18-24 hrs	58	Good Luxuriant	Pink to pinkish red	Complies
<i>Enterobacter aerogenes</i> ATCC 13048	35-37°C / 18-24 hrs	84	Good Luxuriant	Pink to pinkish red	Complies
<i>Salmonella Enteritidis</i> ATCC 13076	35-37°C / 18-24 hrs	75	Good Luxuriant	Colourless to orangish yellow	Complies
<i>Staphylococcus aureus</i> ATCC 25923	35-37°C / 18-24 hrs	≥103	Inhibited	-	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report