



PRODUCT SPECIFICATION SHEET

Vogel-Johnson Agar Medium w/o Tellurite (V.J. Agar) (DM462I)

Intended Use

Vogel-Johnson Agar Medium (DM462I) is recommended for selective isolation of coagulase positive, mannitol fermenting *Staphylococcus aureus* from heavily contaminated foods and clinical specimens in compliance with IP.

Product Summary and Explanation

Staphylococcus aureus, a gram-positive, spherical bacterium, is a common invader of the human skin and mucosa. It is a causative agent of skin and wound infections, urinary tract infections, pneumonia and bacteremia. It is also commonly implicated in food poisoning. It causes spoilage or chemical changes in cosmetic products.⁽¹⁾

In 1955, Zebovitz, Evans, and Niven,⁽²⁾ developed Tellurite-Glycine Agar as a selective plating medium for the quantitative detection of coagulase-positive staphylococci. This medium was modified by Vogel and Johnson in 1960 by the addition of phenol red as a pH indicator and by increasing the mannitol content.⁽³⁾ Vogel-Johnson Agar Medium is prepared in accordance with Indian Pharmacopoeia.⁽⁴⁾ Vogel and Johnson Agar selects and differentiates coagulase-positive staphylococci that ferment mannitol and reduce tellurite.⁽⁵⁾ V.J. Agar is specified in the standard methods for examination of cosmetics,^(1,6) pharmaceutical articles and nutritional supplements.⁽⁴⁾

Principles of the Procedure

Vogel-Johnson Agar Medium contains pancreatic digest of casein which is a source of carbon, nitrogen, vitamins and minerals. Yeast extract supplies B-complex vitamins which stimulate bacterial growth. Mannitol is the carbohydrate. Dibasic potassium phosphate provides buffering to the medium. During the first 24 hours of incubation, contaminating organisms are almost inhibited by tellurite, lithium chloride and high glycine content. Coagulase-positive *staphylococci* reduce potassium tellurite to metallic free tellurium and thus produce black colonies surrounded by yellow zones. This yellow colour is due to phenol red indicator that turns yellow in acidic condition due to the fermentation of mannitol. If mannitol is not fermented, yellow zones are not formed. Also the colour of the medium around the colonies may even be a deeper red than normal due to utilization of the peptones in the medium. Prolonged incubation may result in the growth of black coagulase-negative colonies. Prepared plates of Vogel and Johnson Agar contain 0.2 g/L of potassium tellurite.

Formula / Liter

Ingredients	Gms / Liter
Pancreatic digest of Casein	10.00
Yeast extract	5.00
Mannitol	10.00
Dibasic potassium phosphate	5.00
Lithium chloride	5.00
Glycine	10.00
Phenol red	0.025
Agar	16.00
Final pH: 7.2 ± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	





PRODUCT SPECIFICATION SHEET

Precautions

1. For Laboratory Use only.
2. IRRITANT. Irritating to eyes, respiratory system, and skin.
3. Lithium chloride is harmful. Avoid bodily contact and inhalation of vapours. On contact with skin, wash with plenty of water immediately.

Directions

1. Suspend 61.02 grams of the medium in one liter of distilled water.
2. Heat to boiling, to dissolve the medium completely.
3. Autoclave at 121°C, 15 psi pressure, for 15 minutes / validated cycle.
4. Cool to 45-50°C and add 20 ml of sterile 1% Potassium tellurite solution (MS024).
5. Mix gently and pour into sterile Petri plates.

Quality Control Specifications

Dehydrated Appearance	Light yellow to pink homogeneous free flowing powder
Prepared Medium	Red coloured clear to slightly opalescent gel forms in Petri plates
Reaction of % solution	Not Applicable
Gel Strength	Firm, comparable with 1.6% Agar gel

Growth Promotion Test

Growth promotion is carried out in accordance with Indian Pharmacopoeia.

Expected Cultural Response: Cultural characteristics observed with added 1% Potassium Tellurite solution (MS024), after an incubation at 30-35°C for 18-48 hours.

Sr. No.	Organisms	Inoculum (CFU)	Results to be achieved				
			Growth	Observed Lot Value (CFU)	Recovery	Colour of Colony	Incubation Period
	Test for specified microorganisms						
1.	<i>Staphylococcus aureus</i> ATCC 6538	50-100	good-luxuriant	25-100	≥50 %	black colony surrounded by yellow zone	18 -48 hrs
	Additional Microbiological Testing						
2.	<i>Staphylococcus aureus</i> ATCC 25923	50-100	good-luxuriant	25 -100	≥50 %	black colony surrounded by yellow zone	18 -48 hrs
3.	<i>Staphylococcus epidermidis</i> ATCC 12228	50-100	fair - good	15-40	30 -40 %	translucent to blackish	18 -48 hrs
4.	<i>Proteus mirabilis</i> ATCC 25933	50-100	none-poor	0-10	0 -10 %	yellow	18 -48 hrs





PRODUCT SPECIFICATION SHEET

5.	<i>Escherichia coli</i> ATCC 8739	$\geq 10^3$	inhibited	0	0%	--	≥ 48 hrs
----	-----------------------------------	-------------	-----------	---	----	----	---------------

The organisms listed are the minimum that should be used for quality control testing.

Test Procedure

1. Refer to appropriate references for standard test procedures to obtain isolated colonies from specimens.
2. Incubate plates for 18-48 hours at $35 \pm 2^\circ\text{C}$ in an aerobic atmosphere.

Results

Refer to appropriate references and standard test procedures for interpretation of results.

Storage

Store the sealed bottle containing the dehydrated medium at $10 - 30^\circ\text{C}$. Once opened and recapped, place container in a low humidity environment at the same storage temperature. Protect from moisture and light.

Expiration

Refer to the expiration date stamped on the container. The dehydrated medium should be discarded if not free flowing, or if the appearance has changed from the original color. Expiry applies to medium in its intact container when stored as directed.

Limitations of the Procedure

1. For identification, organisms must be in pure culture. Morphological, biochemical and/or serological tests should be performed for final identification.
2. Consult appropriate texts for detailed information and recommended procedures.

Packaging

Product Name : Vogel-Johnson Agar Medium

Product Code : DM462I

Available Pack sizes : 100gm/500gm

References

1. FDA Bacteriological Analytical Manual, 2005, 18th Ed., AOAC, Washington, D.C.
2. Zebovitz, Evans and Niven. 1955. J. Bacteriol. 70:686.
3. Vogel and Johnson. 1960. Public Health Lab. 18:131.
4. United States Pharmacopeial Convention, Inc. 2008. The United States pharmacopeia 31/The national formulary 26, Supp. 1, 8-1-08, online. United States Pharmacopeial Convention, Inc., Rockville, Md.
5. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams & Wilkins, Baltimore, Md.
6. Curry A. S., Graf J. G. and McEwen G. M., (Eds.), 1993, CTFA Microbiology Guidelines, The Cosmetic, Toiletry and Fragrance Association, Washington, D.C.

Further Information

For further information please contact your local MICROMASTER Representative.








PRODUCT SPECIFICATION SHEET



MICROMASTER LABORATORIES PRIVATE LIMITED
Unit 38/39, Kalpataru Industrial Estate,
Off G.B. Road, Near 'R-Mall', Thane (W) - 400607. M.S. INDIA.
Ph: +91-22-25895505, 4760, 4681. Cell: 9320126789.
Email: micromaster@micromasterlab.com
sales@micromasterlab.com

DM462IPSS, Rev.00, Ver.00/01.02.2016

Prepared By	Checked By	Approved By
 01.01.2018	 01.01.2018	 01.01.2018
Microbiologist	Head Quality Control	Head Quality Assurance

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

