



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: M-FC Agar Base	
Product Code: DM781	Expiry Date : 13/05/2022
Lot Number: 03180517	Release Date: 13/05/2017
Analytical Report Number: QAD/AR-17/05-564	Page No: 1 of 2

DEHYDRATED MEDIUM

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color	Light yellow to greyish yellow	Light yellow to greyish yellow	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies

PREPARED MEDIUM

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color of medium	After Addition of 1% Rosolic Acid : Red coloured	After Addition of 1% Rosolic Acid : Red coloured	Complies
Clarity of medium (After sterilization)	slightly opalescent	slightly opalescent	Complies
pH (After sterilization)	7.20 to 7.60 at 25°C.	7.32	Complies
Gel strength	Firm, comparable with 1.5% Agar gel	Firm, comparable with 1.5% Agar gel	Complies

CULTURAL PARAMETERS:

Organisms	Incubation time/ temp	No. of organisms inoculated	Growth at 35-37°C	Growth at 45.5°C	Colour of colony (on membrane filter)	Remarks
<i>Enterococcus faecalis</i> ATCC 29212	35-37°C 22-24hrs	>=10 ³	Inhibited	Inhibited	-	Complies
<i>Escherichia coli</i> ATCC 25922	35-37°C 22-24hrs	76	Good-luxuriant	Good-luxuriant	light blue	Complies
<i>Salmonella Typhimurium</i> ATCC 14028	35-37°C 22-24hrs	83	Good-luxuriant	Inhibited	pinkish	Complies
<i>Shigella flexneri</i> ATCC 12022	35-37°C 22-24hrs	78	Good-luxuriant	Inhibited	pinkish	Complies

Rev. 00


10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com








CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: M-FC Agar Base	
Product Code: DM781	Expiry Date : 13/05/2022
Lot Number: 03180517	Release Date: 13/05/2017
Analytical Report Number: QAD/AR-17/05-564	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Daniya Kazi) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

