

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Bacitracin Discs (0.04U) (B)	
Product Code: ID001-1vl	Expiry Date : 19/04/2020
Lot Number: 07001016	Release Date: 19/10/2016
Analytical Report Number: QAD/AR-16/10-1346	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 10 mm diameter with printed "B" on each side of the disc.	Filter paper discs of 10 mm diameter with printed "B" on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used with (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Streptococcus pyogenes</i> ATCC19615	Tryptose blood agar Base(DM272)	35-37°C for 18-24 hours.	15-20mm	18	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

