



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Oxidase Discs	
Product Code: ID003-1v1	Expiry Date : 16/02/2019
Lot Number: 00780218	Release Date: 12/02/2018
Analytical Report Number: QAD/AR-18/02-129	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Description	Sterile filter paper discs, impregnated with a solution of Oxidase Reagent	Sterile filter paper discs, impregnated with a solution of Oxidase Reagent	Complies
Oxidase reaction	<i>Pseudomonas aeruginosa</i> ATCC 27853	A dark blue or black purple, positive	Complies
	<i>Escherichia coli</i> ATCC 25922	No colour change, negative	

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

<b>TESTED BY</b> (Deepthi Nair) MICROBIOLOGIST	<b>CHECKED BY</b> (Aditi Modak) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

ID003COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster Laboratories Pvt. Ltd

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwa Estate, Behind R Mall,  
Ghodbunder Road, Thane (W), 400607, Maharashtra, India

Tel : +91-22-25800001, 25800002 | Fax : +91-22-25804760 | sales@micromasterlab.com

