






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Oxidase Discs	
Product Code: ID003-1v1	Expiry Date : 16/02/2019
Lot Number: 00990218	Release Date: 16/02/2018
Analytical Report Number: QAD/AR-18/02-151	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Description	Sterile filter paper discs, impregnated with a solution of Oxidase Reagent	Sterile filter paper discs, impregnated with a solution of Oxidase Reagent	Complies
Oxidase reaction	<i>Pseudomonas aeruginosa</i> ATCC 27853	A dark blue or black purple, positive	Complies
	<i>Escherichia coli</i> ATCC 25922	No colour change, negative	

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Madhura Kadam) MICROBIOLOGIST	<b>CHECKED BY</b> (Aditi Modak) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report