



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Optochin Discs	
Product Code: ID006-1vI	Expiry Date : 11/10/2021
Lot Number: 02190418	Release Date: 10/04/2018
Analytical Report Number: QAD/AR-18/04-352	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 10 mm diameter with printed "OP" on each side of the disc.	Filter paper discs of 10 mm diameter with printed " OP" on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used with (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Streptococcus pneumoniae</i> ATCC 6303	Soyabean casein digest Agar with Sheep Blood (DM247)	35-37°C/ 18-24 hrs	More than or equal to 15mm	16	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Deepti Nair) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report