






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: V- Factor Discs	
Product Code: ID008-1vl	Expiry Date : 21/01/2022
Lot Number: 04250718	Release Date: 24/07/2018
Analytical Report Number: QAD/AR-18/07-677	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Description	Sterile filter paper discs, impregnated with a solution of nicotinamide adenine dinucleotide reagent	Sterile filter paper discs, impregnated with a solution of nicotinamide adenine dinucleotide reagent	Complies
V-Factor Discs	<i>Haemophilus influenzae</i> ATCC 49766	No zone of growth around the disc	Complies
	<i>Haemophilus parainfluenzae</i> ATCC 33392	A clearly defined zone of growth around the disc	
	<i>Haemophilus ducreyi</i> ATCC 700724	No zone of growth around the disc	
	<i>Bordetella pertussis</i> ATCC 9797	A clearly defined zone of growth around the disc	

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

ID008COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,

Ghodbunder Road,Thane (W)-400607,Maharashtra,India

Tel:- +91-22-2589 5505sales@micromasterlab.com