






# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: X+V factor Discs	
Product Code: ID009-1vI	Expiry Date : 21/08/2021
Lot Number: 00930218	Release Date: 05/03/3018
Analytical Report Number: QAD/AR-18/03-204	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Description	Sterile filter paper discs, impregnated with a solution of hemin and nicotinamide adenine dinucleotide reagent	Sterile filter paper discs, impregnated with a solution of hemin and nicotinamide adenine dinucleotide reagent	Complies
X+V Factor Discs	<i>Haemophilus influenzae</i> ATCC 49766	A clearly defined zone of growth around the disc	Complies
	<i>Haemophilus parainfluenzae</i> ATCC 33392	A clearly defined zone of growth around the disc	
	<i>Haemophilus ducreyi</i> ATCC 700724	A clearly defined zone of growth around the disc	
	<i>Bordetella pertussis</i> ATCC 9797	A clearly defined zone of growth around the disc	

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Madhura Kadam) MICROBIOLOGIST	<b>CHECKED BY</b> (Aditi Modak) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

ID009COA,QAD/FR/031,Rev.00/01.01.2018

## Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,  
Ghodbunder Road,Thane (W)-400607,Maharashtra,India

Tel:- +91-22-2589 5505sales@micromasterlab.com