






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Rabbit Plasma (for Coagulase test) (0.1 gm per vial)	
Product Code: IM013R-1vl	Expiry Date : 08/07/2020
Lot Number: 01560317	Release Date: 11/03/2017
Analytical Report Number: QAD/AR-17/03-316	Page No : Page 1 of 1

TESTS CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Sterility Test	Product should pass sterility as per USP36/NF31	Sterility passes	Complies
Coagulase Test	<i>Staphylococcus aureus</i> (25923)	Clot Formation	Clot Formed within 1hour
	<i>Staphylococcus epidermis</i> (12228)	No Clot Formation	No Clot Formed

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Tejal Shinde) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

