






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Kovac's Reagent	
Product Code: IR002-100ml	Expiry Date : 20/01/2019
Lot Number: 00350117	Release Date: 20/01/2017
Analytical Report Number: QAD/AR-17/01-146	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Pale Yellowish colored solution	Pale Yellowish colored solution	Complies
Quality Control strains used	Indole positive: <i>Escherichia coli</i> ATCC 25922: Indicated by a red layer at the top of the broth after the addition of Kovac's reagent.	Indole positive: <i>Escherichia coli</i> ATCC 25922: Indicated by a red layer at the top of the broth after the addition of Kovac's reagent.	Complies
	Indole negative: <i>Enterobacter aerogenes</i> ATCC 13048: Indicated by the lack of color change in the top of the broth after the addition of Kovac's reagent.	Indole negative: <i>Enterobacter aerogenes</i> ATCC 13048: Indicated by the lack of color change in the top of the broth after the addition of Kovac's reagent.	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

