






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Kovac's Reagent	
Product Code: IR002-100ml	Expiry Date : 22/12/2018
Lot Number: 09081216	Release Date: 22/12/2016
Analytical Report Number: QAD/AR-16/12-1670	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Pale Yellowish colored solution	Pale Yellowish colored solution	Complies
Quality Control strains used	<b>Indole positive:</b> <i>Escherichia coli</i> ATCC 25922: Indicated by a red layer at the top of the broth after the addition of Kovac's reagent.	<b>Indole positive:</b> <i>Escherichia coli</i> ATCC 25922: Indicated by a red layer at the top of the broth after the addition of Kovac's reagent.	Complies
	<b>Indole negative:</b> <i>Enterobacter aerogenes</i> ATCC 13048: Indicated by the lack of color change in the top of the broth after the addition of Kovac's reagent.	<b>Indole negative:</b> <i>Enterobacter aerogenes</i> ATCC 13048: Indicated by the lack of color change in the top of the broth after the addition of Kovac's reagent.	Complies

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

