






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Kovac's Reagent	
Product Code: IR002-100ml	Expiry Date : 23/08/2026
Lot Number:05200818	Release Date: 23/08/2018
Analytical Report Number: QAD/AR-18/08-812	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Pale Yellowish colored solution	Pale Yellowish colored solution	Complies
Quality Control strains used	<b>Indole positive:</b> <i>Escherichia coli</i> ATCC 25922: Indicated by a red layer at the top of the broth after the addition of Kovac's reagent.	<b>Indole positive:</b> <i>Escherichia coli</i> ATCC 25922: Indicated by a red layer at the top of the broth after the addition of Kovac's reagent.	Complies
	<b>Indole negative:</b> <i>Enterobacter aerogenes</i> ATCC 13048: Indicated by the lack of color change in the top of the broth after the addition of Kovac's reagent.	<b>Indole negative:</b> <i>Enterobacter aerogenes</i> ATCC 13048: Indicated by the lack of color change in the top of the broth after the addition of Kovac's reagent.	Complies

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Madhura Kadam) <b>MICROBIOLOGIST</b>	<b>CHECKED BY</b> (Aditi Modak) <b>HEAD MICROBIOLOGY</b>	<b>APPROVED BY</b> (Rupali Sabale) <b>HEAD QUALITY ASSURANCE</b>

End Of The Report