



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Barritt's Reagent B	
Product Code: IR033	Expiry Date : 09/04/2026
Lot Number: 02100418	Release Date: 10/04/2018
Analytical Report Number: QAD/AR-18/04-337	Page No : Page 1 of 2

TESTS CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Colourless solution	Colourless solution	Complies
Clarity	Clear without any precipitate	Clear without any precipitate	Complies

CULTURAL PARAMETERS: Biochemical identification was carried out by adding 0.6 ml of Barritt's Reagent A (IR032) and 0.2 ml of Barritt's Reagent B (IR033) in 24-48 hours old cultures grown in MR-VP Medium (DM142)

Organisms	No. of organisms inoculated	Growth	VP Test	Remarks
<i>Enterobacter aerogenes</i> ATCC 13048	90	Luxuriant	Positive (Cherry red colour formation)	Complies
<i>Escherichia coli</i> ATCC 25922	53	Luxuriant	Negative (No red colour formation)	Complies
<i>Klebsiella pneumoniae</i> ATCC 13883	72	Luxuriant	Positive (Cherry red colour formation)	Complies






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Barritt's Reagent B	
Product Code: IR033	Expiry Date : 09/04/2026
Lot Number: 02100418	Release Date: 10/04/2018
Analytical Report Number: QAD/AR-18/04-337	Page No : Page 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report