



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Master Nichrome Loop - 1mm	
Product Code: ML001	Expiry Date: 22/08/2026
Lot Number: G1050818	Release Date: 24/08/2018
Analytical Report Number: QAD/AR-18/08-830	Page No: 1 of 1

Environmental Detail: Temperature: 25.6°C

Humidity: 51%




Test Conducted	Master Loop Diameter(mm)	Master Loop Volume(µL)	Expected Uncertainty
Reading 1	0.08	1.1	+/- 0.02
Reading 2	0.08	1.2	
Reading 3	0.06	1.1	
Reading 4	0.08	1.0	
Reading 5	0.08	1.0	
Reading 6	0.10	1.1	
Reading 7	0.08	1.2	
Reading 8	0.08	1.1	
Reading 9	0.10	1.2	
Reading 10	0.08	1.2	
Average	0.08	1.2	

Traceability:

Instrument Details:		
Name:	ID No	Calibration Due Date
Digital Weighing Balance	MAC-034	---
Weight Box	MI-032	14.05.2018

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report