

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Master Nichrome Loop - 4 mm	
Product Code: ML008	Expiry Date: 13/02/2025
Lot Number: 00960217B	Release Date: 13/02/2017
Analytical Report Number: QAD/AR-17/02-236	Page No: 1 of 1

Environmental Detail: Temperature: 25.6°C

Humidity: 51%

Test Conducted	Master Loop Volume(µL)	Expected Uncertainty
Reading 1	10.1	+/- 0.02
Reading 2	9.9	
Reading 3	10.1	
Reading 4	10.1	
Reading 5	10	
Reading 6	10	
Reading 7	10	
Reading 8	10.2	
Reading 9	10	
Reading 10	10.2	
Average	10.06	
Master Loop Diameter(mm)	4mm	

Traceability:

Instrument Details:		
Name:	ID No	Calibration Due Date
Digital Weighing Balance	MAC-034	---
Weight Box	MI-032	18/05/2017

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

<p>TESTED BY (Divya Sasidharan) MICROBIOLOGIST</p>	<p>CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY</p>	<p>APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE</p>
---	--	--

End Of The Report

