

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Master Nichrome Loop - 4 mm	
Product Code: ML008	Expiry Date: 13/02/2025
Lot Number: 00960217F	Release Date: 13/02/2017
Analytical Report Number: QAD/AR-17/02-240	Page No: 1 of 1

Environmental Detail: Temperature: 25.6°C

Humidity: 51%

Test Conducted	Master Loop Volume(µL)	Expected Uncertainty
Reading 1	9.9	+/- 0.02
Reading 2	10.1	
Reading 3	10	
Reading 4	10.1	
Reading 5	10	
Reading 6	10	
Reading 7	9.9	
Reading 8	10.1	
Reading 9	10.1	
Reading 10	10	
Average	10.02	
Master Loop Diameter(mm)	4mm	

Traceability:

Instrument Details:		
Name:	ID No	Calibration Due Date
Digital Weighing Balance	MAC-034	---
Weight Box	MI-032	18/05/2017

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

Micromaster Laboratories Pvt. Ltd.
 Manufacture & Exporter

10/12/2014

Culturing your new ideas with technology...

QAD/FR/01

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
 Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
 Ph: +91-22-2589 5505, 4760, 4681, 9320126789
 Email: micromaster@micromasterlab.com

