



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name : Vitox Growth Supplement (Twin Pack)	
Product Code: MS035	Expiry Date: 11/01/2022
Lot Number: 03840718	Release Date: 13/07/2018
Analytical Report Number: QAD/AR-18/07-612	Page No: 1 of 2

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Slightly pink in colour	Slightly pink in colour	Complies
Solubility	Rehydrate the contents of 1 vial of Vitox Growth Supplement with 1 vial of Sterile Vitox Hydration Fluid.	Rehydrate the contents of 1 vial of Vitox Growth Supplement with 1 vial of Sterile Vitox Hydration Fluid.	Complies
Sterility testing	A representative sample checked for sterility as per USP33/NF28 passes sterility	A representative sample checked for sterility as per USP33/NF28 passes sterility	Complies
Cultural response	Cultural characteristics observed when used in molten Thayer Martin Medium Base (DM262) / GC Agar Base (DM 116) / Chocolate Agar Base (DM314) along with 250ml of Sterile 2% Hemoglobin Solution.	Cultural characteristics observed when used in molten Thayer Martin Medium Base (DM262) / GC Agar Base (DM 116) / Chocolate Agar Base (DM314) along with 250ml of Sterile 2% Hemoglobin Solution.	Complies

CULTURAL PARAMETERS: Cultural characteristics observed with added sterile lysed blood/Haemoglobin solution (MS088), Vitox Growth Supplement (MS035) and V.C.N. Supplement (MS033) / V.C.N.T. Supplement (MS034) after an incubation at 35-37°C for 18-48 hours

Organisms	Incubation time/ temp	No. of organisms inoculated	No. of organisms recovered	Percentage recovery	Growth	Color of colony	Remarks
<i>Escherichia coli</i> ATCC 25922	35-37°C 18-48 hrs.	$\geq 10^3$	0	0%	Inhibited	--	Complies
<i>Neisseria gonorrhoeae</i> ATCC 19424	35-37°C 18-48 hrs.	84	65	77%	Good-luxuriant	Small, grayish white To Colorless, Mucoid	Complies
<i>Neisseria meningitidis</i> ATCC 13090	35-37°C 18-48 hrs.	81	62	76%	Good-luxuriant	Medium to Large, blue gray, Mucoid	Complies
<i>Proteus mirabilis</i> ATCC 25933	35-37°C 18-48 hrs.	$\geq 10^3$	0	0%	Inhibited	--	Complies

MS035COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road ,Thane (W)-400607,Maharashtra,India

Tel:- +91-22-2589 5505sales@micromasterlab.com






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name : Vitox Growth Supplement (Twin Pack)	
Product Code: MS035	Expiry Date: 11/01/2022
Lot Number: 03840718	Release Date: 13/07/2018
Analytical Report Number: QAD/AR-18/07-612	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

MS035COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road ,Thane (W)-400607,Maharashtra,India

Tel:- +91-22-2589 5505sales@micromasterlab.com