



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Legionella Selective Supplement	
Product Code: MS164	Expiry Date: 02/03/2022
Lot Number: 05140818	Release Date: 29/08/2018
Analytical Report Number: QAD/AR-18/08-852	Page No: 1 of 2

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	White coloured, homogeneous, free flowing powder.	White coloured, homogeneous, free flowing powder.	Complies
Solubility	Contents of 1 vial soluble in 10ml sterile distilled water and forms uniform suspension.	Contents of 1 vial soluble in 10ml sterile distilled water and forms uniform suspension.	Complies
Sterility testing	A representative sample checked for sterility as per USP33/NF28 passes sterility	A representative sample checked for sterility as per USP33/NF28 passes sterility	Complies
Cultural response	Cultural characteristics observed when used in Legionella Agar Base(DM1072A) after an incubation at 35-37°C for 48-72 hours alongwith Legionella Growth Supplement (BCYE) (MS075)	Cultural characteristics observed when used in Legionella Agar Base(DM1072A) after an incubation at 35-37°C for 48-72 hours alongwith Legionella Growth Supplement (BCYE) (MS075)	Complies

CULTURAL PARAMETERS:

Sr. No.	Organisms	No. of organisms inoculated	No. of organisms recovered	Percentage recovery	Growth (MS164)	Colony characteristics	Remarks
1.	<i>Escherichia coli</i> ATCC25922	>10 ³	0	0%	Inhibited	-	Complies
2.	<i>Legionella dumoffii</i> ATCC 33343	78	62	79%	Good-luxuriant	Light blue to grey white colonies	Complies
3.	<i>Legionella pneumophila</i> ATCC 33153	82	60	73%	Good-luxuriant	Light blue to grey white colonies	Complies






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Legionella Selective Supplement	
Product Code: MS164	Expiry Date: 02/03/2022
Lot Number: 05140818	Release Date: 29/08/2018
Analytical Report Number: QAD/AR-18/08-852	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report