



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Novobiocin Selective supplement	
Product Code: MS211	Expiry Date: 25/10/2020
Lot Number: 03540517	Release Date: 10/06/2017
Analytical Report Number: QAD/AR-17/06-626	Page No: 1 of 2

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	White colored, homogeneous, free flowing powder.	White colored, homogeneous, free flowing powder.	Complies
Solubility	Contents of 1 vial soluble in 5ml distilled water	Contents of 1 vial soluble in 5ml distilled water	Complies
Sterility testing	A representative sample checked for sterility as per USP33/NF28 passes sterility	A representative sample checked for sterility as per USP33/NF28 passes sterility	Complies
Cultural response	Cultural characteristics observed when used in Soyabean Bile Broth Base (DM962) Modified Soyabean Bile Broth Base (DM962SO)/ Modified EC Broth Base (DM1649) after an incubation at 35-37°C for 18-48 hours.	Cultural characteristics observed when used in Soyabean Bile Broth Base (DM962) after an incubation at 35-37°C for 18-48 hours.	Complies

CULTURAL PARAMETERS:

Cultural characteristics observed on DM962 after an incubation at 35-37°C or 41.5 °C for 18-48 hours with added Novobiocin Selective Supplement (MS211). Recovery on Sorbitol MacConkey Agar (DM154)

Sr. No.	Organisms	No. of Organisms inoculated	Recovery on DM154	Result
1.	<i>Escherichia coli</i> O157:H7 NCTC12900	81	Good-luxuriant	Complies
2.	<i>Escherichia coli</i> ATCC 25922	76	Good	Complies
3.	<i>Enterococcus faecalis</i> ATCC 29212	>=10 ⁵	Inhibited	Complies
4.	<i>Staphylococcus aureus</i> ATCC 25923	>=10 ⁵	Inhibited	Complies

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall,
Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com








CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Novobiocin Selective supplement	
Product Code: MS211	Expiry Date: 25/10/2020
Lot Number: 03540517	Release Date: 10/06/2017
Analytical Report Number: QAD/AR-17/06-626	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Tejal Shinde) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall,
Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com

