






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Sterile Flexi Loops 2.2 mm	
Product Code: PW035	Expiry Date: 09/06/2022
Lot Number: 03300517	Release Date: 13/06/2017
Analytical Report Number: QAD/AR-17/05-607	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops.	Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops.	Complies
Sterilization	The above batch of Flexi Loops was sterilized by Microtrol Sterilization Services Pvt. Ltd. by treatment with Ethylene oxide to the following parameters: Concentration of Ethylene oxide : NLT 750 gm/M3 Exposure time : 6 hours Humidity : Min. 50% RH Temperature : 50 Deg C	Concentration of Ethylene oxide: 750 gm/M3	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Daniya Kazi) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall,
Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com

