CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Sterile Flexi Loops 2.2 mm	
Product Code: PW035	**Expiry Date: 26/06/2022**
Lot Number: 03420517	**Release Date: 14/08/2017**
Analytical Report Number: QAD/AR-17/07-823	**Page No: 1 of 1**

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops.	Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops.	Complies
Sterilization	The above batch of Flexi Loops was sterilized by Microtrol Sterilization Services Pvt. Ltd. by treatment with Ethylene oxide to the following parameters: Concentration of Ethylene oxide : NLT 750 gm/M3 Exposure time : 6 hours Humidity : Min. 50% RH Temperature : 50 Deg C	Concentration of Ethylene oxide: 750 gm/M3	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY **(Shubhangi Kachare)** **MICROBIOLOGIST**	**CHECKED BY** **(Tejal Shinde)** **HEAD MICROBIOLOGY**	**APPROVED BY** **(Rupali Sabale)** **HEAD QUALITY ASSURANCE**

End Of The Report

Micromaster Laboratories Pvt. Ltd

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall, Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com