

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY






Product Name: Gram's Iodine	
Product Code: SI016	Expiry Date :25/04/2025
Lot Number: 02770417	Release Date: 25/04/2017
Analytical Report Number: QAD/AR-17/04-462	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Reddish Brown colored solution	Reddish Brown colored solution	Complies
Staining of Blood Smear	Gram Staining procedure followed using the reagent should show appropriate results	Gram Staining procedure followed using the reagent showed appropriate results	Complies
Quality Control strains used	<i>Staphylococcus aureus</i> ATCC 25923	Gram-positive cocci	Complies
	<i>Escherichia coli</i> ATCC 25922	Gram-negative rods	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Daniya Kazi) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

