

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY






Product Name: Methylene Blue (Aqueous)	
Product Code: SI023	Expiry Date : 25/04/2025
Lot Number: 02780417	Release Date: 25/04/2017
Analytical Report Number: QAD/AR-17/04-464	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Deep Blue Colored Solution	Deep Blue Colored Solution	Complies
Staining of Bacterial Suspension	ZNCF Staining procedure followed using the reagent should show appropriate results	ZNCF Staining procedure followed using the reagent showed appropriate results	Complies
Quality Control strains used	<i>Mycobacterium tuberculosis H37Ra</i> ATCC 25177	Pinkish Red colored Rods observed	Complies
	<i>Bacillus subtilis</i> ATCC 6633	Blue colored Rods observed	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Daniya Kazi) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

