






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Gram's Staining Kit (4 x 250ml)	
Product Code: SK001-Kit	Expiry Date : 19/07/2026
Lot Number: 04270718	Release Date: 20/07/2018
Analytical Report Number: QAD/AR-18/07-652	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Kit Contains the following <ul style="list-style-type: none">Crystal VioletGram's DecolorizerGram's IodineSafranin - 0.5% w/v	Kit Contains the following <ul style="list-style-type: none">Crystal VioletGram's DecolorizerGram's IodineSafranin - 0.5% w/v	Complies
Staining of Bacterial Suspension	Gram's Staining procedure followed using the reagent should show appropriate results	Gram's Staining procedure followed using the reagent showed appropriate results	Complies
Quality Control strains used	<i>Staphylococcus aureus</i> ATCC 25923	Gram-positive cocci	Complies
	<i>Escherichia coli</i> ATCC 25922	Gram-negative rods	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Pradnya Nadivkar) MICROBIOLOGIST	CHECKED BY (Rupali Sabale) HEAD MICROBIOLOGY	APPROVED BY (Nilesh Shinde) HEAD QUALITY ASSURANCE

End Of The Report

SK001COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road ,Thane (W)-400607,Maharashtra,India

Tel:- +91-22-2589 5505sales@micromasterlab.com