



PRODUCT SPECIFICATION SHEET

Transport Swabs with Amies Charcoal Medium (Pack of 50) (TM001)

Intended Use

Transport swab with Amies Charcoal Medium is recommended for recovery of aerobic, anaerobic and fastidious bacteria from throat, vaginal and wound specimens.

Product Summary and Explanation

Transport media are described as chemically defined, semisolid, non-nutritive, phosphate buffered media that provide a reduced environment. Transport media should be able to hamper self-destructive enzymatic reactions within the cells and in addition, must inhibit toxic oxidation reactions. These media are formulated to retain the viability of microorganisms without significant increase in growth. Amies⁽¹⁾ modified the formulation of Stuart's Transport Medium^(2, 3, 4) by replacing glycerophosphate with an inorganic phosphate buffer and adding charcoal to the medium. This modified medium yielded a higher percentage of positive cultures than the transport medium of Stuart. Transport Medium Amies is recommended for throat, vaginal and wound samples. Amies media are especially suited for specimens containing *Neisseria gonorrhoeae*.

Principles of the Procedure

Amies Transport Medium w/ Charcoal provides a reduced environment due to the presence of sodium thioglycollate and small amount of agar. Charcoal helps to neutralize materials that are toxic to sensitive pathogens like *Neisseria gonorrhoeae*. The medium contains potassium chloride, calcium chloride, magnesium chloride and sodium chloride which provide essential ions that help maintain osmotic balance while controlling permeability of bacterial cells. Monopotassium phosphate and disodium phosphate provide buffering capabilities.

Formula / Liter

Ingredients	Gms / Liter
Sodium chloride	3.00
Potassium chloride	0.20
Calcium chloride	0.10
Magnesium chloride	0.10
Monopotassium phosphate	0.02
Disodium phosphate	1.15
Sodium thioglycollate	1.00
Charcoal	10.00
Agar	4.00
Final pH: 7.2 ± 0.2 at 25°C	
Formula may be adjusted and/or supplemented as required to meet performance specifications	

Precautions

For In-Vitro Diagnostic Use Only.

Quality Control Specifications

Prepared Medium	Black coloured opaque gel forms in tubes as butts
Reaction of 2% Solution	pH: 7.2 ± 0.2 at 25°C





PRODUCT SPECIFICATION SHEET

Media Per tube	3ml
----------------	-----

Sterility test

Passes release criteria

Expected Cultural Response: Cultural characteristics observed when subcultured on Soyabean Casein Digest Agar (DM247) after an incubation at 35-37°C for 18-24 hours.

Sr. No.	Organisms	Results to be achieved	
		Inoculum (CFU)	Growth
1.	<i>Escherichia coli</i> ATCC 25922	50 - 100	good-luxuriant
2.	<i>Klebsiella pneumonia</i> ATCC 13883	50 - 100	good-luxuriant
3.	<i>Neisseria meningitidis</i> ATCC 13090	50 - 100	good-luxuriant
4.	<i>Pseudomonas aeruginosa</i> ATCC 27853	50 - 100	good-luxuriant
5.	<i>Salmonella Typhi</i> ATCC 6539	50 - 100	good-luxuriant
6.	<i>Shigella flexneri</i> ATCC 12022	50 - 100	good-luxuriant
7.	<i>Staphylococcus aureus</i> ATCC 25923	50 - 100	good-luxuriant
8.	<i>Vibrio cholerae</i> ATCC 15748	50 - 100	good-luxuriant

The organisms listed are the minimum that should be used for quality control testing.

Test Procedure

1. Refer to appropriate references for specific procedures.
2. Obtain specimen with sterile swab sterile cotton-tipped swabs or wooden sticks. Insert specimen swab(s) into the upper third of the medium in the transport container. Cut with sterile scissors or break-off the protruding portion of the swab stick.
3. Tightly screw the lid on the bottle or vial or plug the tube with cotton, due to which the swab is forced to the bottom of the medium.
4. Label the bottle or vial and send to the laboratory with minimum delay. Specimens may be refrigerated until ready for shipment. DO NOT FREEZE.
5. Submit to laboratory within 24 hours for culture and analysis.

Results

1. Refer to appropriate references and standard procedures for interpretation of results.
2. Survival of bacteria in a transport medium depends on many factors including the type and concentration of bacteria in the specimen, the formulation of the transport medium, the temperature and duration of transport and inoculation to appropriate culture media within 24 hours.
3. Optimal growth and typical morphology can only be expected following direct inoculation and appropriate cultivation.

Storage

Store at 2-8°C. The product retains potency until the expiration date shown on the label when stored properly under ideal storage conditions.

Expiration

Refer to the expiration date stamped on the label.





PRODUCT SPECIFICATION SHEET

Limitations of the Procedure

1. Specimens taken from transport media will not exhibit the optimal or comparative growth as expected from direct inoculation and cultivation. These media do, however, provide an adequate degree of preservation for those specimens which cannot be forwarded immediately to the laboratory for prompt evaluation.
2. Viability of cells will diminish over time and some degree of multiplication or growth of contaminants can occur during prolonged periods of transit. This is particularly true of faecal specimens that contain substantial numbers of coliform organisms.
3. The condition of the specimen received by the laboratory for culture is a significant variable in recovery and final identification of the suspect pathogen. An unsatisfactory specimen (overgrown by contaminants, containing nonviable organisms, or having the number of pathogens greatly diminished) can lead to erroneous or inconclusive results.

Packaging

Product Name : Transport swab with Amies Charcoal Medium

Product Code : TM001

Available Pack sizes : 50 TB

References

1. Amies C.R., 1967, Can. J. Public Health, 58:296
2. Stuart R.D., 1946, J. Path. Bact., 58:343.
3. Stuart R.D., 1959, Pub. Hlth. Rep., 74:431.
4. Stuart R.D., Toshach S.R. and Patsula T.M., 1954, Can. J. Pub. Hlth., 45:75.

Further Information

For further information please contact your local MICROMASTER Representative.



MICROMASTER LABORATORIES PRIVATE LIMITED


TM001PSS, QAD/FR/024, Rev.00/01.01.2018

Unit 38/39, Kalpataru Industrial Estate,

Off G.B. Road, Near 'R-Mall', Thane (W) - 400607. M.S. INDIA.

Ph: +91-22-25895505, 4760, 4681. Cell: 9320126789.

Email: micromaster@micromasterlab.com

Prepared By	Checked By	Approved By
 01.01.2018	 01.01.2018	 01.01.2018
Microbiologist	Head Quality Control	Head Quality Assurance





PRODUCT SPECIFICATION SHEET

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

