



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Transport Swabs with Stuart Medium	
Product Code: TM004-20TB	Expiry Date: 10/10/2019
Lot Number: 01980418	Release Date: 08/04/2018
Analytical Report Number: QAD/AR-18/04-330	Page No: 1 of 2

PREPARED MEDIA			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Sterile transport swabs with Stuart medium	Sterile transport swabs with Stuart medium	
Color of Final Medium	Whitish to light blue colored medium	Whitish to light blue colored medium	Complies
Clarity of Final Medium	Slightly opalescent butt with upper 10% or less portion blue on standing	Slightly opalescent butt with upper 10% or less portion blue on standing	Complies
pH at 25°C	7.20 to 7.60 at 25°C	7.36	Complies
Gel Strength	Semisolid, comparable with 0.3% Agar gel	Semisolid, comparable with 0.3% Agar gel	Complies
Quantity of Medium	2ml	2ml	Complies
PREINCUBATION CHECK OF PLATE			
Pre Incubation Check	No growth should be observed after incubation at 30-35 °C for 24 - 48 hrs	No growth observed after incubation at 30-35 °C for 24 - 48 hrs	Complies

CULTURAL PARAMETERS: Viability of following organisms was established for a period of 48 hours. Organisms grew luxuriantly when inoculated on Chocolate Agar(M103) and incubated at 35 - 37°C for 18-24 hours.

Organisms	Incubation time/ temp	No. of organisms inoculated	Growth	Remarks
<i>Neisseria gonorrhoeae</i> ATCC 19424	35-37°C / 18-24 hrs	77	Luxuriant	Complies
<i>Streptococcus pneumoniae</i> ATCC 6303	35-37°C / 18-24 hrs	72	Luxuriant	Complies
<i>Haemophilus influenzae</i> ATCC 35056	35-37°C / 18-24 hrs	75	Luxuriant	Complies






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Transport Swabs with Stuart Medium	
Product Code: TM004-20TB	Expiry Date: 10/10/2019
Lot Number: 01980418	Release Date: 08/04/2018
Analytical Report Number: QAD/AR-18/04-330	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Deepti Nair) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report