

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Sterile Swab with 0.85% Saline (Pack of 25) (TM005)	
Product Code: TM005	Expiry Date: 21/03/2019
Lot Number: 06450917	Release Date: 22/09/2017
Analytical Report Number: QAD/AR-17/09-1039	Page No: 1 of 2

PREPARED MEDIA			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Sterile transport swab 0.85% Saline	Sterile transport swab with 0.85% Saline	Complies
Colour of Final Medium	Colorless solution	Colorless solution	Complies
Clarity of Final Medium	Colorless solution	Colorless solution	Complies
pH at 25°C	6.80 $\pm$ 0.50	6.80	Complies
Gel Strength	Not Applicable	Not Applicable	Complies
Quantity of Medium	2 ml	2 ml	Complies
PREINCUBATION CHECK OF PLATE			
Pre Incubation Check	No growth should be observed after incubation at 30-35 °C for 24 - 48 hrs	No growth observed after incubation at 30-35 °C for 24 - 48 hrs	Complies

**CULTURAL PARAMETERS:** Viability of following organisms was established for a period of 24 hours. Organisms grew luxuriantly when recovered on Soyabean Casein Digest Agar (DM247) and incubated at 35-37°C for 18-24 hours.

Organisms	Incubation time/ temp	No. of organisms inoculated	Growth	Remarks
Escherichia coli ATCC 25922	35-37°C / 18-24 hrs	86	good-luxuriant	Complies
Staphylococcus aureus ATCC 25923	35-37°C / 18-24 hrs	82	good-luxuriant	Complies
Pseudomonas aeruginosa ATCC 27853	35-37°C / 18-24 hrs	87	good-luxuriant	Complies

Rev. 00

10/12/2014

QAD/FR/01

**Micromaster Laboratories Pvt. Ltd**

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall,  
Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com








# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Sterile Swab with 0.85% Saline (Pack of 25) (TM005)	
Product Code: TM005	Expiry Date: 21/03/2019
Lot Number: 06450917	Release Date: 22/09/2017
Analytical Report Number: QAD/AR-17/09-1039	Page No: 2 of 2

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Tejal Shinde) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

**Micromaster Laboratories Pvt. Ltd**

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall,  
Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com

