






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Viro-Master Transport Medium</b>	
<b>Product Code: TM006-25TB</b>	<b>Expiry Date: 14/11/2019</b>
<b>Lot Number: 02730518</b>	<b>Release Date: 15/05/2018</b>
<b>Analytical Report Number: QAD/AR-18/05-456</b>	<b>Page No: 1 of 1</b>

<b>PREPARED MEDIA</b>			
<b>TEST CONDUCTED</b>	<b>SPECIFICATION</b>	<b>OBSERVATIONS</b>	<b>REMARKS</b>
Appearance	Sterile transport swab with Viral transport medium	Sterile transport swab with Viral transport medium	
Color of Final Medium	Light Pink colored Medium	Light Pink colored Medium	Complies
Clarity of Final Medium	Clear Solution	Clear Solution	Complies
Quantity of Medium	3ml	3ml	Complies
<b>PREINCUBATION CHECK OF PLATE</b>			
Pre Incubation Check	No growth should be observed after incubation at 30-35 °C for 24 - 48 hrs	No growth observed after incubation at 30-35 °C for 24 - 48 hrs	Complies

### **Final Results: COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY (Madhura Kadam) MICROBIOLOGIST</b>	<b>CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY</b>	<b>APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE</b>

End Of The Report